

General Engineering Laboratories, Inc., Charleston, SC.
 2040 Savage Rd
 Charleston SC 29407

Chain of Custody/Analysis Request

COC/Lab Request #:
 12-1508

Page 1 of 1

Client Contact:				Lab Agreement # : 126310011				Site Name: Los Alamos National Laboratory															
Project Number :				WSP-GENINORG	WSP-GrossA/B	WSP-Met+B+SN+SR+U	WSP-NH3+NO3/NO2+PO4	WSP-RAD	WSP-TKN+TOC													Rad Screening Info:	
Analysis Turnaround Time:																						Yes, Below Background	
24 Hour - <input type="checkbox"/> Other - <input type="checkbox"/>																							
7 Day - <input type="checkbox"/>																							
14 Day - <input type="checkbox"/>																							
21 Day - <input type="checkbox"/>																							
28 Day - <input checked="" type="checkbox"/>				Special Instructions:																			
Field Sample ID	Sample Date	Sample Time	Sample Matrix																				
CASA-12-21643	Aug 17 2012	11:01	W	1	1			1	1														
CASA-12-21647	Aug 17 2012	11:01	W	1	1	1																	

Special Instructions:

Relinquished by: <i>Meliss Montoy</i>	<i>[Signature]</i>	Date/Time: 8/20/12 3:00	Received by:
Relinquished by:		Date/Time:	Received by:
Relinquished by:		Date/Time:	Received by:

SAMPLE COLLECTION LOG/FIELD CHAIN OF CUSTODY

EVENT ID: 3953 EVENT NAME: Sandia (Chromium Monitoring)
 Q4 Watershed Sampling
 SAMPLE ID: CASA-12-21647 WORK ORDER: NA

	<u>AS PLANNED</u>	<u>AS COLLECTED</u>		<u>AS PLANNED</u>	<u>AS COLLECTED</u>
DATE COLLECTED (MM/DD/YYYY):		08/17/12	FIELD MATRIX:	WG	OK
TIME COLLECTED (HH:MM):		1101	MEDIA:	UA	OK
PRS ID:		OK	SAMPLE TECH CODE:	UA	GSP
LOCATION ID: R-11			FIELD PREP:	F	OK
LOCATION TYPE: MON			FIELD QC TYPE:	REG	
PORT: SINGLE COMPLETION			SAMPLE USAGE:	INV	

PRIORITY	ORDER	CONTAINER	#	PRESERVATIVE	COLLECTED Y/N	SPECIAL INSTRUCTIONS
NA	WSP-GENINORG	1 LITER POLY	1	ICE	Y	NA
	WSP-Met+B+SN+SR+U	1 LITER POLY	1	HNO3		
	WSP-NH3+NO3/NO2+PO4	500 ML AMBER GLASS	1	H2SO4		

SAMPLE COMMENTS:

LOCATION COMMENTS:

FIELD PARAMETERS:

Dissolved Oxygen _____ mg/L Oxidation-Reduction Potential _____ MV pH _____ SU
 Specific Conductance _____ uS/cm Temperature _____ deg C Turbidity _____ NTU

COLLECTED BY (PRINT) *D Fellenz*

RELINQUISHED BY (Printed Name) <i>D Woody</i> (Signature) <i>David Woody</i>	Date/Time 08/17/12 1155	RECEIVED BY (Printed Name) <i>S. Sherwood</i> (Signature) <i>Sheri Sherwood</i>	Date/Time 08/17/12 1155
RELINQUISHED BY (Printed Name) (Signature)	Date/Time	RECEIVED BY (Printed Name) (Signature)	Date/Time

Report Date 07/30/2012

See
CASA-12-21643

Data Validation Report

Chain Of Custody No. 12-1508

1. Distribution Of Samples In EDD.

SDG	Analytical Method	Regular Samples	Field Duplicates	Trip Blanks	Field Blanks	Equipment Blanks
309911	EPA:120.1		1			
309911	EPA:150.1		1			
309911	EPA:160.1		1			
309911	EPA:245.2		1			
309911	EPA:300.0		1			
309911	EPA:310.1		1			
309911	EPA:350.1		1			
309911	EPA:351.2		1			
309911	EPA:353.2		1			
309911	EPA:365.4		1			
309911	EPA:900		1			
309911	EPA:901.1		1			
309911	EPA:905.0		1			
309911	HASL-300:AM-241		1			
309911	HASL-300:ISOPU		1			
309911	HASL-300:ISOU		1			
309911	SM:A2340B		1			
309911	SW-846:6010B		1			
309911	SW-846:6020		1			
309911	SW-846:6850		1			
309911	SW-846:9060		1			

SDG	Analytical Method	Analysis Lot ID	Prep Lot ID	Regular Samples	Field Duplicates	Trip Blanks	Field Blanks	Equipment Blanks	Method Blanks	Matrix Spikes	Matrix Spike Dups
309911	EPA:120.1	1242995	1242995		1						
309911	EPA:150.1	1241830	1241830		1						
309911	EPA:160.1	1240523	1240523		1					1	
309911	EPA:245.2	1243838	1243833		1					1	1
309911	EPA:300.0	1241176	1241176		1					1	
309911	EPA:310.1	1242061	1242061		1					1	2
309911	EPA:350.1	1240583	1240582		1					1	1
309911	EPA:351.2	1240586	1240585		1					1	2
309911	EPA:353.2	1240584	1240584		1					1	
309911	EPA:365.4	1240581	1240580		1					1	2
309911	EPA:900	1239941	1239941		1					1	1
309911	EPA:901.1	1240464	1240464		1					1	
309911	EPA:905.0	1239939	1239939		1					1	1
309911	HASL-300:AM-241	1239554	1239554		1					1	
309911	HASL-300:ISOPU	1239555	1239555		1					1	
309911	HASL-300:ISOU	1239557	1239557		1					1	
309911	SM:A2340B	1244727	1244727		1						
309911	SW-846:6010B	1240142	1240141		1					1	1
309911	SW-846:6020	1240139	1240137		1					1	1
309911	SW-846:6850	1241188	1241187		1					1	1
309911	SW-846:9060	1241231	1241231		1					1	

2. Distribution Of Analytes In EDD.

Analytical Spikes	Post-Digestion Spikes	Lab Control Samples	Lab Control Sample Dups	Blank Spikes	Blank Spike Dups	Lab Duplicates	Storage Blanks	Preparation Blanks	Reagent Blanks
		1				1			
		1				2			
		1				1			
		1				1			
		1				1			
		1				2			
		1				1			
		1				2			
		1				1			
		1				2			
		1				1			
		1				1			
		1				1			
		1				1			
		1				1			
		1				1			
		1				1			
		1				1			
		1				1			
		1				2			

Analytical Method	Method Category	Field Sample ID	Lab Sample ID	Sample Purpose	Target Analytes	Surrogates	Spikes	TICS
EPA:120.1	GENERAL CHEMISTRY	CAMO-12-21742	1202731308	DUP		1	0	0
EPA:120.1	GENERAL CHEMISTRY	CASA-12-21647	309911002	REG		1	0	0
EPA:120.1	GENERAL CHEMISTRY	LCS	1202731309	LCS		0	0	1
EPA:150.1	GENERAL CHEMISTRY	CAMO-12-21742	1202728425	DUP		1	0	0
EPA:150.1	GENERAL CHEMISTRY	CASA-12-21647	309911002	REG		1	0	0
EPA:150.1	GENERAL CHEMISTRY	LCS	1202728426	LCS		0	0	1
EPA:150.1	GENERAL CHEMISTRY	WTRO-12-23145	1202728424	DUP		1	0	0
EPA:160.1	GENERAL CHEMISTRY	CAMO-12-21742	1202725255	DUP		1	0	0
EPA:160.1	GENERAL CHEMISTRY	CASA-12-21647	309911002	REG		1	0	0
EPA:160.1	GENERAL CHEMISTRY	LCS	1202725256	LCS		0	0	1
EPA:160.1	GENERAL CHEMISTRY	MB	1202725254	MB		1	0	0
EPA:245.2	INORGANIC	CAMO-12-21742	1202733108	DUP		1	0	0
EPA:245.2	INORGANIC	CAMO-12-21742	1202733109	MS		0	0	1
EPA:245.2	INORGANIC	CASA-12-21647	309911002	REG		1	0	0
EPA:245.2	INORGANIC	LCS	1202733107	LCS		0	0	1
EPA:245.2	INORGANIC	MB	1202733106	MB		1	0	0
EPA:300.0	GENERAL CHEMISTRY	CASA-12-21647	309911002	REG		4	0	0
EPA:300.0	GENERAL CHEMISTRY	CASA-12-21772	1202726830	DUP		4	0	0
EPA:300.0	GENERAL CHEMISTRY	LCS	1202726832	LCS		0	0	4
EPA:300.0	GENERAL CHEMISTRY	MB	1202726829	MB		4	0	0
EPA:310.1	GENERAL CHEMISTRY	CASA-12-21647	309911002	REG		2	0	0
EPA:310.1	GENERAL CHEMISTRY	CASA-12-21775	1202729070	DUP		3	0	0
EPA:310.1	GENERAL CHEMISTRY	CASA-12-21775	1202729071	MS		0	0	1
EPA:310.1	GENERAL CHEMISTRY	LCS	1202729069	LCS		0	0	1
EPA:310.1	GENERAL CHEMISTRY	MB	1202729072	MB		3	0	0
EPA:310.1	GENERAL CHEMISTRY	WTRO-12-23137	1202729076	DUP		3	0	0
EPA:310.1	GENERAL CHEMISTRY	WTRO-12-23137	1202729077	MS		0	0	1
EPA:350.1	GENERAL CHEMISTRY	CAMO-12-21796	1202725418	DUP		1	0	0
EPA:350.1	GENERAL CHEMISTRY	CAMO-12-21796	1202725420	MS		0	0	1
EPA:350.1	GENERAL CHEMISTRY	CAMO-12-21796	1202725422	MSD		0	0	1
EPA:350.1	GENERAL CHEMISTRY	CASA-12-21647	309911002	REG		1	0	0
EPA:350.1	GENERAL CHEMISTRY	LCS	1202725424	LCS		0	0	1
EPA:350.1	GENERAL CHEMISTRY	MB	1202725417	MB		1	0	0
EPA:351.2	GENERAL CHEMISTRY	CAMO-12-21734	1202730363	DUP		1	0	0
EPA:351.2	GENERAL CHEMISTRY	CAMO-12-21734	1202730364	MS		0	0	1
EPA:351.2	GENERAL CHEMISTRY	CAMO-12-21734	1202730365	MSD		0	0	1
EPA:351.2	GENERAL CHEMISTRY	CASA-12-21643	309911001	REG		1	0	0
EPA:351.2	GENERAL CHEMISTRY	LCS	1202725437	LCS		0	0	1
EPA:351.2	GENERAL CHEMISTRY	MB	1202725433	MB		1	0	0
EPA:351.2	GENERAL CHEMISTRY	SWWS46-12-22930	1202725434	DUP		1	0	0
EPA:351.2	GENERAL CHEMISTRY	SWWS46-12-22930	1202725435	MS		0	0	1
EPA:351.2	GENERAL CHEMISTRY	SWWS46-12-22930	1202725436	MSD		0	0	1
EPA:353.2	GENERAL CHEMISTRY	CAMO-12-21796	1202725426	DUP		1	0	0
EPA:353.2	GENERAL CHEMISTRY	CASA-12-21647	309911002	REG		1	0	0
EPA:353.2	GENERAL CHEMISTRY	LCS	1202725432	LCS		0	0	1
EPA:353.2	GENERAL CHEMISTRY	MB	1202725425	MB		1	0	0
EPA:365.4	GENERAL CHEMISTRY	CAMO-12-21743	1202725413	DUP		1	0	0
EPA:365.4	GENERAL CHEMISTRY	CAMO-12-21743	1202725414	MS		0	0	1
EPA:365.4	GENERAL CHEMISTRY	CAMO-12-21743	1202725415	MSD		0	0	1
EPA:365.4	GENERAL CHEMISTRY	CAMO-12-21744	1202730366	DUP		1	0	0
EPA:365.4	GENERAL CHEMISTRY	CAMO-12-21744	1202730367	MS		0	0	1
EPA:365.4	GENERAL CHEMISTRY	CAMO-12-21744	1202730368	MSD		0	0	1

EPA:365.4	GENERAL CHEMISTRY	CASA-12-21647	309911002	REG	1	0	0	0
EPA:365.4	GENERAL CHEMISTRY	LCS	1202725416	LCS	0	0	1	0
EPA:365.4	GENERAL CHEMISTRY	MB	1202725412	MB	1	0	0	0
EPA:900	RAD	CAMO-12-21736	1202723793	DUP	2	0	0	0
EPA:900	RAD	CAMO-12-21736	1202723796	MS	0	0	2	0
EPA:900	RAD	CAMO-12-21736	1202723797	MSD	0	0	2	0
EPA:900	RAD	CASA-12-21643	309911001	REG	2	0	0	0
EPA:900	RAD	LCS	1202723798	LCS	0	0	2	0
EPA:900	RAD	MB	1202723792	MB	2	0	0	0
EPA:901.1	RAD	CASA-12-21643	309911001	REG	5	0	0	0
EPA:901.1	RAD	LCS	1202725098	LCS	0	0	3	0
EPA:901.1	RAD	MB	1202725096	MB	17	0	0	0
EPA:901.1	RAD	WST03-12-23161	1202725097	DUP	17	0	0	0
EPA:905.0	RAD	CASA-12-21643	1202723780	DUP	1	0	0	0
EPA:905.0	RAD	CASA-12-21643	1202723781	MS	0	0	1	0
EPA:905.0	RAD	CASA-12-21643	309911001	REG	1	0	0	0
EPA:905.0	RAD	LCS	1202723782	LCS	0	0	1	0
EPA:905.0	RAD	MB	1202723779	MB	1	0	0	0
HASL-300:AM-241	RAD	CAMO-12-21738	1202722700	DUP	1	0	0	0
HASL-300:AM-241	RAD	CASA-12-21643	309911001	REG	1	0	0	0
HASL-300:AM-241	RAD	LCS	1202722701	LCS	0	0	1	0
HASL-300:AM-241	RAD	MB	1202722699	MB	1	0	0	0
HASL-300:ISOPU	RAD	CAMO-12-21738	1202722703	DUP	2	0	0	0
HASL-300:ISOPU	RAD	CASA-12-21643	309911001	REG	2	0	0	0
HASL-300:ISOPU	RAD	LCS	1202722704	LCS	0	0	1	0
HASL-300:ISOPU	RAD	MB	1202722702	MB	2	0	0	0
HASL-300:ISOU	RAD	CAMO-12-21738	1202722706	DUP	3	0	0	0
HASL-300:ISOU	RAD	CASA-12-21643	309911001	REG	3	0	0	0
HASL-300:ISOU	RAD	LCS	1202722707	LCS	0	0	1	0
HASL-300:ISOU	RAD	MB	1202722705	MB	3	0	0	0
SM:A2340B	INORGANIC	CASA-12-21647	309911002	REG	1	0	0	0
SW-846:6010B	INORGANIC	Bay-39	1202724302	DUP	22	0	0	0
SW-846:6010B	INORGANIC	Bay-39	1202724303	MS	0	0	22	0
SW-846:6010B	INORGANIC	CASA-12-21647	309911002	REG	17	0	0	0
SW-846:6010B	INORGANIC	LCS	1202724301	LCS	0	0	22	0
SW-846:6010B	INORGANIC	MB	1202724300	MB	22	0	0	0
SW-846:6020	INORGANIC	Bay-39	1202724294	DUP	13	0	0	0
SW-846:6020	INORGANIC	Bay-39	1202724295	MS	0	0	13	0
SW-846:6020	INORGANIC	CASA-12-21647	309911002	REG	11	0	0	0
SW-846:6020	INORGANIC	LCS	1202724293	LCS	0	0	13	0
SW-846:6020	INORGANIC	MB	1202724292	MB	13	0	0	0
SW-846:6850	LCMS/MS PERCHLORATE	CASA-12-21647	309911002	REG	1	0	0	0
SW-846:6850	LCMS/MS PERCHLORATE	CASA-12-21773	1202726854	MS	0	0	1	0
SW-846:6850	LCMS/MS PERCHLORATE	CASA-12-21773	1202726855	MSD	0	0	1	0
SW-846:6850	LCMS/MS PERCHLORATE	LCS	1202726853	LCS	0	0	1	0
SW-846:6850	LCMS/MS PERCHLORATE	MB	1202726852	MB	1	0	0	0
SW-846:9060	GENERAL CHEMISTRY	CAMO-12-21779	1202726979	DUP	1	0	0	0
SW-846:9060	GENERAL CHEMISTRY	CASA-12-21643	1202726978	DUP	1	0	0	0
SW-846:9060	GENERAL CHEMISTRY	CASA-12-21643	309911001	REG	1	0	0	0
SW-846:9060	GENERAL CHEMISTRY	LCS	1202726982	LCS	0	0	1	0
SW-846:9060	GENERAL CHEMISTRY	MB	1202726977	MB	1	0	0	0

3. Are any analytes missing?

No.

4. Were any holding times exceeded?

No.

5. Any contaminants in blanks?

Field	Lab	Type Of	Analytical	Sample	Parameter	Lab	Lab		Lab
Sample ID	Sample ID	Blank	Method	Matrix	Name	Result	Qualifier	Units	Detection Limit
MB	1202724300	METHOD BLANK	SW-846:6010B	W	Potassium	-153		ug/L	150
MB	1202725412	METHOD BLANK	EPA:365.4	W	Total Phosphate as Phosphorus	0.0454	J	mg/L	0.05
MB	1202725417	METHOD BLANK	EPA:350.1	W	Ammonia as Nitrogen	0.0359	J	mg/L	0.05
MB	1202733106	METHOD BLANK	EPA:245.2	W	Mercury	-0.085	J	ug/L	0.2

Any samples affected by the presence of contaminants in blanks?

Field	Blank Field	Blank Lab	Blank	Analytical	Parameter		Blank	Sample	Lab	Detect	
Sample ID	Sample ID	Sample ID	Type	Method	Name	Units	Result	Result	Qualifier	Limit	Detected
CASA-12-21647	MB	1202733106	METHOD BLANK	EPA:245.2	Mercury	ug/L	-0.085	0.067	U	0.2	N
CASA-12-21647	MB	1202725417	METHOD BLANK	EPA:350.1	Ammonia as Nitrogen	mg/L	0.0359	0.124		0.05	Y
CASA-12-21647	MB	1202725412	METHOD BLANK	EPA:365.4	Total Phosphate as Phosphorus	mg/L	0.0454	0.0349	J	0.05	Y
CASA-12-21647	MB	1202724300	METHOD BLANK	SW-846:6010B	Potassium	ug/L	-153	1540		150	Y

6. Any surrogate recoveries outside the control limits?

No.

7. Any MS/MSD recoveries or RPDs outside the control limits?

Field	Matrix	Matrix	Analytical	Parameter	Analysis	Analysis	Sample	MS %	MSD %	Upper	Lower
Sample ID	Spike ID	Spike Dup ID	Method	Name	Lot ID	Date	Matrix	Recvry	Recvry	Limit	Limit
CAMO-12-21734	1202730364	1202730365	EPA:351.2	Total Kjeldahl Nitrogen	1240585	9/5/2012	W	85.7	90	110	90

8. Any LCS/LCSD or BS/BSD recoveries or RPDs outside the control limits?

No.

9. Any Field Duplicate RPDs outside the desired limits?

No.

10. Any Lab Duplicate RPDs outside the desired limits?

No.

11. Any required reporting limits exceeded?

No.

Correction Factor (ND)	Correction Factor (J)	Use Factors
5		Y
5		Y
5		Y
5		Y

Rejection Limit	RPD	RPD Limit
10	4.42	20

12. Additional Validator's Comments.

None.

13. Display Flagged Data.

Location ID	Chain Of Custody No	Field Sample ID	Sample Purpose	Analysis Type Code	Analytical Suite	Analytical Method	Parameter Name	Lab Qualifier	Validation Qualifier	Validation Reason Codes	Detected
R-11	12-1508	CASA-12-21643	REG	INIT	RAD	HASL-300:AM-241	Americium-241	U	U	R5	N
R-11	12-1508	CASA-12-21643	REG	INIT	RAD	EPA:901.1	Cesium-137	U	U	R5	N
R-11	12-1508	CASA-12-21643	REG	INIT	RAD	EPA:901.1	Cobalt-60	U	U	R5	N
R-11	12-1508	CASA-12-21643	REG	INIT	RAD	EPA:900	Gross alpha	U	U	R5	N
R-11	12-1508	CASA-12-21643	REG	INIT	RAD	EPA:900	Gross beta	U	U	R5	N
R-11	12-1508	CASA-12-21643	REG	INIT	RAD	EPA:901.1	Neptunium-237	U	U	R5	N
R-11	12-1508	CASA-12-21643	REG	INIT	RAD	HASL-300:ISOPU	Plutonium-238	U	U	R5	N
R-11	12-1508	CASA-12-21643	REG	INIT	RAD	HASL-300:ISOPU	Plutonium-239/240	U	U	R5	N
R-11	12-1508	CASA-12-21643	REG	INIT	RAD	EPA:901.1	Potassium-40	U	U	R5	N
R-11	12-1508	CASA-12-21643	REG	INIT	RAD	EPA:901.1	Sodium-22	U	U	R5	N
R-11	12-1508	CASA-12-21643	REG	INIT	RAD	EPA:905.0	Strontium-90	U	U	R5	N
R-11	12-1508	CASA-12-21643	REG	INIT	RAD	HASL-300:ISOU	Uranium-235/236	U	U	R5	N
R-11	12-1508	CASA-12-21647	REG	INIT	GENERAL CHEMISTRY	EPA:350.1	Ammonia as Nitrogen		U	I4	N
R-11	12-1508	CASA-12-21647	REG	INIT	GENERAL CHEMISTRY	EPA:365.4	Total Phosphate as Phosphorus	J	U	I4	N

Reason Code Description
 I4 the sample result is <=5x the concentration of related analyte in the method blank.

J_LAB The analytical laboratory qualified the detected result as estimated (J) because the result was less the PQL but greater than the MDL

NQ The analytical laboratory did not qualify the analyte as not detected and/or any other standard qualifire. The analyte is detected in the sample.

R5 Analyte is not detected because the amount reported is less than the MDC.

U_LAB The analytical laboratory qualified the analyte as not detected.

14. Useable Result Count.

Field	Location	Sample	Analytical	No. Unuseable	Total No. Of
Sample ID	ID	Purpose	Method	Records	Records
CASA-12-21643	R-11	REG	EPA:351.2	0	1
CASA-12-21643	R-11	REG	EPA:900	0	2
CASA-12-21643	R-11	REG	EPA:901.1	0	5
CASA-12-21643	R-11	REG	EPA:905.0	0	1
CASA-12-21643	R-11	REG	HASL-300:AM-241	0	1
CASA-12-21643	R-11	REG	HASL-300:ISOPU	0	2
CASA-12-21643	R-11	REG	HASL-300:ISOU	0	3
CASA-12-21643	R-11	REG	SW-846:9060	0	1
CASA-12-21647	R-11	REG	EPA:120.1	0	1
CASA-12-21647	R-11	REG	EPA:150.1	0	1
CASA-12-21647	R-11	REG	EPA:160.1	0	1
CASA-12-21647	R-11	REG	EPA:245.2	0	1
CASA-12-21647	R-11	REG	EPA:300.0	0	4
CASA-12-21647	R-11	REG	EPA:310.1	0	2
CASA-12-21647	R-11	REG	EPA:350.1	0	1
CASA-12-21647	R-11	REG	EPA:353.2	0	1
CASA-12-21647	R-11	REG	EPA:365.4	0	1

Lab Result	Lab Units	Report Result	Report Units	Report MDA	Report Uncertainty	Lab Matrix	Sample Date	Percent Moisture	Analysis Lot ID	Validation Status Code	Use Flag
0.00254	pCi/L	0.00254	pCi/L	0.0348	0.00568	W	8/17/2012		1239554	VAL	Y
1.44	pCi/L	1.44	pCi/L	7.55	1.86	W	8/17/2012		1240464	VAL	Y
-2.91	pCi/L	-2.91	pCi/L	7.34	2.18	W	8/17/2012		1240464	VAL	Y
-0.134	pCi/L	-0.134	pCi/L	2.15	0.376	W	8/17/2012		1239941	VAL	Y
1.7	pCi/L	1.7	pCi/L	2.66	0.823	W	8/17/2012		1239941	VAL	Y
-6.01	pCi/L	-6.01	pCi/L	10.6	3.29	W	8/17/2012		1240464	VAL	Y
0	pCi/L	0	pCi/L	0.0171	0.00359	W	8/17/2012		1239555	VAL	Y
0.00761	pCi/L	0.00761	pCi/L	0.0306	0.00567	W	8/17/2012		1239555	VAL	Y
-3.04	pCi/L	-3.04	pCi/L	76.7	17	W	8/17/2012		1240464	VAL	Y
-0.956	pCi/L	-0.956	pCi/L	6.84	1.78	W	8/17/2012		1240464	VAL	Y
0.149	pCi/L	0.149	pCi/L	0.467	0.135	W	8/17/2012		1239939	VAL	Y
0.0136	pCi/L	0.0136	pCi/L	0.0475	0.00833	W	8/17/2012		1239557	VAL	Y
0.124	mg/L	0.124	mg/L			W	8/17/2012		1240583	VAL	Y
0.0349	mg/L	0.0349	mg/L			W	8/17/2012		1240581	VAL	Y

Data Validation Report for:

Chain Of Custody No. 12-1508

CASA-12-21647	R-11	REG	SM:A2340B	0	1
CASA-12-21647	R-11	REG	SW-846:6010B	0	17
CASA-12-21647	R-11	REG	SW-846:6020	0	11
CASA-12-21647	R-11	REG	SW-846:6850	0	1



August 29, 2012

www.gel.com

Keith Greene
Los Alamos National Laboratory
PO Box 1663
TA-03, SM271, Drop Pt. 02U, Rm111
Los Alamos, New Mexico 87545

Re: LANL-WQH Water Samples
Work Order: 309911
SDG: 12-1508

Dear Keith Greene:

GEL Laboratories, LLC (GEL) appreciates the opportunity to provide the following analytical results for the sample(s) we received on August 21, 2012, and analyzed for General Chemistry, Metals, Perchlorates by LCMSMS and Radiochemistry. This original data report has been prepared and reviewed in accordance with GEL's standard operating procedures.

Our policy is to provide high quality, personalized analytical services to enable you to meet your analytical needs on time every time. We trust that you will find everything in order and to your satisfaction. If you have any questions, please do not hesitate to call me at (843) 556-8171, ext. 4485.

Sincerely,

Hope Taylor for
Valerie Davis
Project Manager

Purchase Order: 63641-10
Chain of Custody: 12-1508
Enclosures



ARS International (63641-10)
LANL-WQH Water Samples
Work Order #: 309911
SDG: 12-1508

Table of Contents

Case Narrative.....	1
Chain of Custody and Supporting Documentation.....	4
Data Review Qualifier Flag Definition Sheet.....	10
Perchlorates by LCMSMS Analysis.....	12
Case Narrative.....	13
Sample Data Summary.....	19
Quality Control Summary.....	21
Quality Control Data.....	24
Miscellaneous.....	30
Metals Analysis.....	32
Case Narrative.....	33
Sample Data Summary.....	39
Quality Control Summary.....	43
Miscellaneous.....	57
General Chem Analysis.....	59
Case Narrative.....	60
Sample Data Summary.....	88
Quality Control Summary.....	93
Miscellaneous.....	99
Radioloigcal Analysis.....	103

Sample Data Summary.....	115
Quality Control Data.....	118

Case Narrative

**Case Narrative for
ARS International (63641-10)
LANL-WQH Water Samples
Workorder #: 309911
SDG # : 12-1508**

August 29, 2012

Laboratory Identification:

GEL Laboratories LLC
2040 Savage Road
Charleston, South Carolina 29407
(843) 556-8171

Summary

Sample receipt The samples arrived at GEL Laboratories LLC, Charleston, South Carolina on August 21, 2012 for analysis. Please see attached email for discrepancies. The samples were screened according to GEL Standard Operating Procedure. All sample containers arrived without any visible signs of tampering or breakage. Containers were checked for pH, where appropriate, and matched the preservative as documented on the accompanying chain of custody. Shipping container temperature was within specification (0 - 6C). Shipping container temperatures were checked, documented, and within specifications. There are no additional comments concerning sample receipt.

Sample Identification The laboratory received the following samples:

<u>Laboratory ID</u>	<u>Client ID</u>
309911001	CASA-12-21643
309911002	CASA-12-21647


Case Narrative

Sample analyses were conducted using methodology as outlined in GEL Laboratories, LLC (GEL) Standard Operating Procedures. Any technical or administrative problems during analysis, data review, and reduction are contained in the analytical case narratives in the enclosed data package.

Data Package

The enclosed data package contains the following sections: Case Narrative, Chain of Custody, Cooler Receipt Checklist, Data Package Qualifier Definitions and data from the following fractions: General Chemistry, Metals, Perchlorates by LCMSMS and Radiochemistry.

I certify that this data report is in compliance with the terms and conditions of the subcontract and task order, both technically and for completeness, for other than the conditions detailed in the attached case narrative.



Hope Taylor for
Valerie Davis
Project Manager

List of current GEL Certifications as of 29 August 2012

State	Certification
Arizona	AZ0766
Arkansas	88-0651
CLIA	42D0904046
California NELAP	01151CA
Colorado	SC00012
Connecticut	PH-0169
Delaware	SC00012
DoD ELAP A2LA ISO 17025	2567.01
Florida NELAP	E87156
Foreign Soils Permit	P330-09-00191
Georgia	SC00012
Georgia SDWA	967
Hawaii	SC00012
Idaho	SC00012
Illinois NELAP	200029
Indiana	C-SC-01
Kansas NELAP	E-10332
Kentucky	90129
Louisiana NELAP	03046 (AI33904)
Louisiana SDWA	LA120008
Maryland	270
Massachusetts	M-SC012
Mississippi	SC00012
Nevada	SC000122011-1
New Hampshire NELAP	2054
New Jersey NELAP	SC002
New Mexico	SC00012
New York NELAP	11501
North Carolina	233
North Carolina SDWA	45709
Oklahoma	9904
Pennsylvania NELAP	68-00485
South Carolina Chemistry	10120001
South Carolina Radiochemi	10120002
Tennessee	TN 02934
Texas NELAP	T104704235-12-7
Utah NELAP	SC00012
Vermont	VT87156
Virginia NELAP	460202
Washington	C780
Wisconsin	999887790

Chain of Custody and Supporting Documentation

General Engineering Laboratories, Inc., Charleston, SC.
 2040 Savage Rd
 Charleston SC 29407

Chain of Custody/Analysis Request

309911

COC/Lab Request #:
12-1508

Page 1 of 1

Client Contact:

Lab Agreement # : 126310011

Project Number :

Analysis Turnaround Time:

- 24 Hour - Other -
- 7 Day -
- 14 Day -
- 21 Day -
- 28 Day -

Site Name: Los Alamos National Laboratory

Rad Screening Info:

Yes, Below Background

Field Sample ID

CASA-12-21643
 CASA-12-21647

Sample Date
 Aug 17 2012
 Aug 17 2012

Sample Time
 11:01
 11:01

Sample Matrix
 W
 W

WSP-GENINORG

1

WSP-GrossA/B

1

WSP-Met+B+SN+SR+U

1

WSP-NH3+NO3/NO2+PO4

1

WSP-RAD

1

WSP-TKN+TOC

1

Special Instructions:

Special Instructions:

Relinquished by:

Relinquished by:

Relinquished by:

Date/Time:

8/20/12 3:00

Date/Time:

Date/Time:

Received by:

P. Nert

Received by:

8-21-12 09:40

Received by:



SAMPLE RECEIPT & REVIEW FORM

Client: LANL		SDG/AR/COC/Work Order:12-1508	
Received By: Patricia Dent		Date Received: AUGUST 21, 2012	
Suspected Hazard Information	Yes	No	*If Net Counts > 100cpm on samples not marked "radioactive", contact the Radiation Safety Group for further investigation.
COC/Samples marked as radioactive?		X	Maximum Net Counts Observed* (Observed Counts - Area Background Counts): 0CPM
Classified Radioactive II or III by RSO?		X	If yes, Were swipes taken of sample containers < action levels?
COC/Samples marked containing PCBs?		X	
Shipped as a DOT Hazardous?		X	Hazard Class Shipped: UN#:
Samples identified as Foreign Soil?		X	

Sample Receipt Criteria	Yes	NA	No	Comments/Qualifiers (Required for Non-Conforming Items)
1 Shipping containers received intact and sealed?	X			Circle Applicable: Seals broken Damaged container Leaking container Other (describe)
2 Samples requiring cold preservation within (0 ≤ 6 deg. C)?*	X			Preservation Method: Ice bags Blue ice Dry ice None Other (describe) *all temperatures are recorded in Celsius 2,3C
2a Daily check performed and passed on IR temperature gun?	X			Temperature Device Serial #: Secondary Temperature Device Serial # (If Applicable): 61524646
3 Chain of custody documents included with shipment?	X			
4 Sample containers intact and sealed?	X			Circle Applicable: Seals broken Damaged container Leaking container Other (describe)
5 Samples requiring chemical preservation at proper pH?	X			Sample ID's, containers affected and observed pH: If Preservation added, Lot#
6 VOA vials free of headspace (defined as < 6mm bubble)?		X		Sample ID's and containers affected:
7 Are Encore containers present?			X	(If yes, immediately deliver to Volatiles laboratory)
8 Samples received within holding time?	X			ID's and tests affected:
9 Sample ID's on COC match ID's on bottles?	X			Sample ID's and containers affected:
10 Date & time on COC match date & time on bottles?	X			Sample ID's affected:
11 Number of containers received match number indicated on COC?			X	Sample ID'S affected: CASA-12-21643 for Gross A/B, WSP-RAD container not received
12 Are sample containers identifiable as GEL provided?			X	CLIENT
13 COC form is properly signed in relinquished/received sections?	X			
14 Carrier and tracking number.	X			Circle Applicable: FedEx Air FedEx Ground UPS Field Services Courier Other 7209 7856 9157 2C 7209 7856 9168 3C 7209 7856 9179 3C

Comments (Use Continuation Form if needed):

Good afternoon all listed below are today's Issues

**RN#12-1508 lab did not receive containers for GROSS A/B,WSP-RAD SAMPLE
CASA-12-21643**

**RN#2012-2209 lab received (1) container with double label for all samples, COC
indicates (2) containers each.**

RN#2012-2212 lab did not receive any containers for WTRO-12-22695,23117,23081,22587.

Thanks! Pat Dent

--

Patricia Dent
Project Manager Assistant
GEL Laboratories, LLC
2040 Savage Rd.
Charleston, S.C. 29407
Main: 843-556-8171 Ext 4264
Fax: 843-766-1178
Email: pad@gel.com
Web: www.gel.com

Subject: RE: ISSUES FROM TODAY 08/21/2012
From: "Greene, Keith R" <kgreene@lanl.gov>
Date: 8/21/2012 4:39 PM
To: Pat Dent <Pat.Dent@gel.com>

Will reciev all missing samples tomorrow we left them on the counter yesterday OOPS

From: Pat Dent [mailto:Pat.Dent@gel.com]
Sent: Tuesday, August 21, 2012 2:39 PM
To: Greene, Keith R
Cc: LANL@amrad.com; team.davis
Subject: ISSUES FROM TODAY 08/21/2012

Good afternoon all listed below are today's Issues

RN#12-1508 lab did not receive containers for GROSS A/B,WSP-RAD SAMPLE CASA-12-21643

RN#2012-2209 lab received (1) container with double label for all samples, COC indicates (2) containers each.

RN#2012-2212 lab did not receive any containers for WTRO-12-22695,23117,23081,22587.

Thanks! Pat Dent

--
Patricia Dent
Project Manager Assistant
GEL Laboratories, LLC
2040 Savage Rd.
Charleston, S.C. 29407
Main: 843-556-8171 Ext 4264
Fax: 843-766-1178
Email: pad@gel.com
Web: www.gel.com

ORIGIN ID: SAFA (505) 665-9966
KEITH GREENE
LOS ALAMOS NATL LAB
TA00 BLDG 1237 DPU 03

SHIP DATE: 20AUG12
ACTWGT: 23.0 LB MAN
CAD: 0014176/CAFE2511

LOS ALAMOS, NM 87545
UNITED STATES US

BILL SENDER

ORIGIN ID: SAFA (505) 665-9966
KEITH GREENE
LOS ALAMOS NATL LAB
TA00 BLDG 1237 DPU 03

SHIP DATE: 20AUG12
ACTWGT: 42.0 LB MAN
CAD: 0014176/CAFE2511

LOS ALAMOS, NM 87545
UNITED STATES US

BILL SENDER

TO VALERIE DAVIS
GENERAL ENGINEERING LAB
2040 SAVAGE RD



TO VALERIE DAVIS
GENERAL ENGINEERING LAB
2040 SAVAGE RD

CHARLESTON SC 29407
(843) 556-8171
REF: MROA002068B0

CHARLESTON SC 29407
(843) 556-8171
REF: MR1A015AGWKO

3c

2c



TRK# 7209 7856 9168
0201

TUE - 21 AUG A1
PRIORITY OVERNIGHT

TRK# 7209 7856 9157
0201

TUE - 21 AUG A1
PRIORITY OVERNIGHT


XX CHSA

29407
SC-US CHS

XX CHSA

29407
SC-US CHS

Part # 156148-434 RITZ 10/11



Part # 156148-434 RITZ 10/11




ORIGIN ID: SAFA (505) 665-9966
KEITH GREENE
LOS ALAMOS NATL LAB
TA00 BLDG 1237 DPU 03
LOS ALAMOS, NM 87545
UNITED STATES US

SHIP DATE: 20AUG12
ACTWGT: 23.0 LB MAN
CAD: 0014176/CAFE2511
BILL SENDER

TO VALERIE DAVIS
GENERAL ENGINEERING LAB
2040 SAVAGE RD

CHARLESTON SC 29407
(843) 556-8171
REF: MR0A00EA5BDO

3



TRK# 7209 7856 9179
0201

TUE - 21 AUG A1
PRIORITY OVERNIGHT

XX CHSA

29407
SC-US CHS

590C1/A270/18BC

Data Review Qualifier Flag Definition Sheet

Data Review Qualifier Definitions

Qualifier	Explanation
*	A quality control analyte recovery is outside of specified acceptance criteria
**	Analyte is a surrogate compound
<	Result is less than value reported
>	Result is greater than value reported
^	RPD of sample and duplicate evaluated using +/-RL. Concentrations are <5X the RL
A	The TIC is a suspected aldol-condensation product
B	Target analyte was detected in the associated blank
B	Metals-Either presence of analyte detected in the associated blank, or MDL/IDL < sample value < PQL
BD	Results are either below the MDC or tracer recovery is low
C	Analyte has been confirmed by GC/MS analysis
D	Results are reported from a diluted aliquot of the sample
d	5-day BOD-The 2:1 depletion requirement was not met for this sample
E	Organics-Concentration of the target analyte exceeds the instrument calibration range
E	Metals-%difference of sample and SD is >10%. Sample concentration must meet flagging criteria
H	Analytical holding time was exceeded
h	Preparation or preservation holding time was exceeded
J	Value is estimated
N	Metals-The Matrix spike sample recovery is not within specified control limits
N	Organics-Presumptive evidence based on mass spectral library search to make a tentative identification of the analyte (TIC). Quantitation is based on nearest internal standard response factor
N/A	Spike recovery limits do not apply. Sample concentration exceeds spike concentration by 4X or more
ND	Analyte concentration is not detected above the reporting limit
UI	Gamma Spectroscopy-Uncertain identification
X	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier
Y	QC Samples were not spiked with this compound
Z	Paint Filter Test-Particulates passed through the filter, however no free liquids were observed.

Perchlorates by LCMSMS Analysis

Case Narrative

**Perchlorate by LC/MSMS
ARS International (ARSL)
SDG 12-1508**

Method/Analysis Information

Procedure: **Definitive Low Level Perchlorate Analysis Utilizing Liquid Chromatography/Mass Spectrometry/Mass Spectrometry (LC/MS/MS) by EPA Method 6850 Modified (6850M)**

Analytical Method: SW846 6850 Modified
Prep Method: SW846 6850 Modified
Analytical Batch Number: 1241188
Prep Batch Number: 1241187

Sample Analysis

Sample ID	Client ID
309911002	CASA-12-21647
1202726856	Interference Check Sample (ICS)
1202726852	Method Blank (MB)
1202726853	Laboratory Control Sample (LCS)
1202726854	310180007(CASA-12-21773) Matrix Spike (MS)
1202726855	310180007(CASA-12-21773) Matrix Spike Duplicate (MSD)

The samples in this SDG were analyzed on an "as received" basis.

Preparation/Analytical Method Verification

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-OA-E-067 REV# 9.

Calibration Information

Initial Calibration

All initial calibration requirements have been met for this SDG.

Due to software constraints, all Initial Calibration Blanks must be designated as IPB001.

ICV Requirements

All associated initial calibration verification standards (ICV) met the acceptance criteria.

CCB Requirements

All continuing calibration blanks (CCB) bracketing the analyses associated with this batch were within acceptance criteria.

CCV Requirements

All continuing calibration checks (CCV) requirements were met by all bracketing CCV standards.

Low Level Standard (CRI) Requirements

All low level calibration verification (CRI) requirements were met by all bracketing CRI standards.

Quality Control (QC) Information**Method Blank (MB) Statement**

The MB analyzed with this SDG met the acceptance criteria.

Interference Check Sample (ICS)

The interference check sample (ICS) met all recovery acceptance criteria.

Laboratory Control Sample (LCS) Recovery

The LCS spike recoveries met the acceptance limits.

QC Sample Designation

Client sample 310180007 (CASA-12-21773) from SDG 12-1513 was chosen for matrix spike and matrix spike duplicate analysis.

Matrix Spike (MS) Recovery Statement

The MS recoveries were within the established acceptance limits.

Matrix Spike Duplicate (MSD) Recovery Statement

High recoveries for Perchlorate and Perchlorate-101 were observed in MSD (1202726855). The recoveries were 179% and 168%, respectively. The acceptance range is 75-125%. The high recoveries may be the result of the background concentration present in the parent sample, 310180007 (CASA-12-21773). Recoveries in the LCS (1202726853) were acceptable. Data are reported with the appropriate DER.

MS/MSD Relative Percent Difference (RPD) Statement

The RPD(s) between the MS and MSD met the acceptance limits.

Retention Time Standard Area Acceptance

The retention time standard areas were within the required acceptance criteria for all samples and QC.

Retention Time

During the analysis of Perchlorate by LC/MS/MS, retention time shifts are commonly observed. These retention time shifts, which are caused by fouling of the column by the sample matrices, are problematic when the retention time is used as one of the criterion for confirmation. To overcome this problem, a known amount of O(18) labeled Perchlorate was added to each sample as a retention time standard.

The presence of Perchlorate was confirmed by the relative retention time (RRT) of the Perchlorate peak and the O(18) standard. A RRT window of 0.98 to 1.02, as required by Method 332.0, has been used.

In addition to the isotopic ratio, the presence of Perchlorate in the samples associated with this data package have been confirmed using the relative retention criteria stated above, not the absolute retention time.

Technical Information

Holding Time Specifications

All samples in this SDG in this analytical batch met the specified holding time. GEL assigns holding times based on the associated methodology, which assigns the date and time from sample collection of sample receipt. Those holding times expressed in hours are calculated in the AlphaLIMS system. Those holding times expressed as days expire at midnight on the day of expiration.

Preparation/Analytical Method Verification

All procedures were performed as stated in the SOP.

Sample Dilutions

The samples in this SDG did not require dilutions.

Sample Re-extraction/Re-analysis

Sample 309911002 (CASA-12-21647) was re-analyzed to confirm the potential of carryover from the previous sample. The re-analysis data are reported.

Miscellaneous Information

Data Exception (DER) Documentation

Data Exception Report 1116450 was generated for this SDG.

High recoveries for Perchlorate and Perchlorate-101 were observed in MSD (1202726855). The recoveries were 179% and 168%, respectively. The acceptance range is 75-125%. The high recoveries may be the result of the background concentration present in the parent sample, 310180007 (CASA-12-21773). Recoveries in the LCS (1202726853) were acceptable.

Manual Integrations

Some initial calibration standards, continuing calibration standards, and/or samples may require manual integrations due to software limitations.

Method Comments

The samples in this SDG were not originally analyzed using EPA Method 314.0.

Additional Comments

The Perchlorate Isotope Ratio on the Form I may differ slightly from the ratio on the corresponding raw data due to rounding rules and/or significant figures or due to software limitations when there are manual integrations, dilutions or other factors. The ratio value of the Form I is the correct value.

The retention time marker, Perchlorate-O (18), is added to all samples, instrument blanks, and standards prior to injection. It is used to verify the retention time of Perchlorate and Perchlorate-101 and to insure an accurate injection occurred.

Due to various anions affecting the recovery of Perchlorate-O (18) and not Perchlorate and Perchlorate-101, the calibration curves of Perchlorate and Perchlorate-101 are not internally corrected for using Perchlorate-O (18). They are external calibrations.

Perchlorate Isotope Ratio

The Perchlorate isotope ratio met acceptance criteria for all samples and QC samples. Please see the isotope ratio criteria in the Miscellaneous Section.

System Configuration

The laboratory utilizes a Waters LC 2795 liquid chromatography instrument for Perchlorate analysis. It is coupled with either a Micromass Quattro Micro Mass Spectrometer/ Mass Spectrometer, or a Micromass Quattro Ultima Mass Spectrometer/ Mass Spectrometer. Each being designated as LCMSMS #1 and LCMSMS #2, respectively. It is fitted with an electrospray probe that is operated in the negative electrospray ionization mode for Perchlorate analysis.

The laboratory may also utilize an Agilent 1100 liquid chromatography instrument for Perchlorate analysis. It is coupled with an Applied Biosystems 4000 Mass Spectrometer/ Mass Spectrometer, designated as LCMSMS #3 or LCMSMS #4. It is also fitted with an electrospray probe that is operated in the negative electrospray ionization mode for Perchlorate analysis.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. An electronic signature page inserted after the case narrative will include the data validator's signature and title. The signature page also includes the data qualifiers used in the fractional package.

Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Chromatographic Columns

Chromatographic separation of Perchlorate is accomplished through analysis on the following anion column:

Dionex: IonPac AG-16 2 x 50 mm.

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Qualifier Definition Report for

ARSL001 ARS International (63641-10)

Client SDG: 12-1508 GEL Work Order: 309911

The Qualifiers in this report are defined as follows:

- * A quality control analyte recovery is outside of specified acceptance criteria
- ** Analyte is a surrogate compound
- U Analyte was analyzed for, but not detected above the MDL, MDA, or LOD.
- DL Indicates that sample is diluted.
- RA Indicates that sample is re-analyzed without re-extraction.
- RE Indicates that sample is re-extracted.

Review/Validation

GEL requires all analytical data to be verified by a qualified data reviewer. In addition, all CLP-like deliverables receive a third level review of the fractional data package.

The following data validator verified the information presented in this data report:

Signature: 

Name: Michael Penny

Date: 07 SEP 2012

Title: Group Leader

Sample Data Summary

Perchlorate Analysis Data Sheet

Lab Name: GEL Laboratories LLC

Client Sample No.

CASA-12-21647Lab Code: GELDate Received: 21-AUG-12Instrument: LCMSMSGEL Job No (SDG): 12-1508Method: SW846 6850 ModifiedGEL Sample ID: 309911002Matrix: WATERDate Filtered: 30-AUG-12Extraction Batch ID: 1241187Injection Volume (uL): 20Extraction Type: Filter/DAISample Volume/Weight: 10.0 mL

%Solids: .

Concentrated Extract Volume: 10.0

CAS No.	Analyte^	MDL	RL	Conc*	Units	Q	Dilution Factor	Date Analyzed	GEL File ID
14797-73-0	Perchlorate	.05	.2	0.899	ug/L		1	31-AUG-12 13:05	per0831014a
	Perchlorate Isotope Ratio			3.35			1	31-AUG-12 13:05	per0831014a
14797-73-0	Perchlorate-101	.05	.2	0.881	ug/L		1	31-AUG-12 13:05	per0831014a
	Perchlorate-O(18)			0.564	ug/L		1	31-AUG-12 13:05	per0831014a

^ When the analyte name is Perchlorate Isotope Ratio the concentration is a unitless value calculated from the ratio of Perchlorate peak area to Perchlorate-101 peak area. The Perchlorate-101 and isotopic ratio results are provided for qualitative purposes only. The results are used to verify the presence and quantitation of Perchlorate.

*Concentration =

$$\text{Instrument Value} \times \frac{\text{Concentrated Extract Volume}}{\text{Aliquot}} \times \frac{1}{\% \text{Solids}}$$

Quality Control Summary

Perchlorate Laboratory Control Sample

Lab Name: General Engineering Laboratories

Lab Code: GEL

GEL Job No. (SDG): 12-1508

Extract Batch Code: 1241187

Date Filtered: 30-AUG-12

Matrix: WATER

Sample ID: 1202726853

Analyte^	True	Found	Units	%Rec	Q	Control Limits
Perchlorate	0.200	.207	ug/L	104		85 - 115
Perchlorate Isotope Ratio		3.31				-
Perchlorate-101	0.200	.205	ug/L	103		85 - 115
Perchlorate-O(18)		.507	ug/L			-

^ When the analyte name is Perchlorate Isotope Ratio the concentration is a unitless value calculated from the ratio of Perchlorate peak area to Perchlorate-101 peak area. The Perchlorate-101 and isotopic ratio results are provided for qualitative purposes only. The results are used to verify the presence and quantitation of Perchlorate.

Perchlorate Spike/Spike Duplicate Summary

Lab Name: General Engineering Laboratories

Lab Code: GEL

GEL Job No (SDG): 12-1508

Extract Batch Code: 1241187

Date Extracted: 30-AUG-12

GEL MS/PS ID: 1202726854

Client ID: CASA-12-21773

GEL MSD/PSD ID: 1202726855

QC Type: MS

Compound^	Spike Added	Sample Conc	Units	MS Conc	MS Rec #	MSD Conc	MSD Rec #	RPD #	RPD Limit	Recovery Limit
Perchlorate	0.200	0.531	ug/L	0.723	95.8	.888	179 *	20.6	30	75 - 125
Perchlorate Isotope Ratio	0	3.27		3.32		3.36		1.06		-
Perchlorate-101	0.200	0.532	ug/L	0.713	90.6	.867	168 *	19.5	30	75 - 125
Perchlorate-O(18)	0	0.526	ug/L	0.528		.532		.903		-

^ When the analyte name is Perchlorate Isotope Ratio the concentration is a unitless value calculated from the ratio of Perchlorate peak area to Perchlorate-101 peak area. The Perchlorate-101 and isotopic ratio results are provided for qualitative purposes only. The results are used to verify the presence and quantitation of Perchlorate.

Quality Control Data

Perchlorate Analysis Data Sheet

Lab Name: GEL Laboratories LLCLab Code: GELInstrument: LCMSMSMethod: EPA 6850 ModifiedMatrix: WATERExtraction Batch ID: 1241187Extraction Type: Filter/DAISample Volume/Weight: 10.0 mLConcentrated Extract Volume: 10.0

Client Sample No.

MBDate Received: 30-AUG-12GEL Job No (SDG): 12-1508GEL Sample ID: 1202726852Date Filtered: 30-AUG-12Injection Volume (uL): 20

%Solids: .

CAS No.	Analyte^	MDL	RL	Conc*	Units	Q	Dilution Factor	Date Analyzed	GEL File ID
14797-73-0	Perchlorate	.05	.2	0.050	ug/L	U	1	30-AUG-12 16:54	per0830012a
	Perchlorate Isotope Ratio						1	30-AUG-12 16:54	per0830012a
14797-73-0	Perchlorate-101	.05	.2	0.050	ug/L	U	1	30-AUG-12 16:54	per0830012a
	Perchlorate-O(18)			0.482	ug/L		1	30-AUG-12 16:54	per0830012a

^ When the analyte name is Perchlorate Isotope Ratio the concentration is a unitless value calculated from the ratio of Perchlorate peak area to Perchlorate-101 peak area. The Perchlorate-101 and isotopic ratio results are provided for qualitative purposes only. The results are used to verify the presence and quantitation of Perchlorate.

*Concentration =

$$\text{Instrument Value} \times \frac{\text{Concentrated Extract Volume}}{\text{Aliquot}} \times \frac{1}{\% \text{Solids}}$$

Perchlorate Analysis Data Sheet

Lab Name: GEL Laboratories LLC

Lab Code: GEL

Instrument: LCMSMS

Method: EPA 6850 Modified

Matrix: WATER

Extraction Batch ID: 1241187

Extraction Type: Filter/DAI

Sample Volume/Weight: 10.0 mL

Concentrated Extract Volume: 10.0

Client Sample No.

LCS

Date Received: 30-AUG-12

GEL Job No (SDG): 12-1508

GEL Sample ID: 1202726853

Date Filtered: 30-AUG-12

Injection Volume (uL): 20

%Solids: .

CAS No.	Analyte^	MDL	RL	Conc*	Units	Q	Dilution Factor	Date Analyzed	GEL File ID
14797-73-0	Perchlorate	.05	.2	0.207	ug/L		1	30-AUG-12 17:02	per0830013a
	Perchlorate Isotope Ratio			3.31			1	30-AUG-12 17:02	per0830013a
14797-73-0	Perchlorate-101	.05	.2	0.205	ug/L		1	30-AUG-12 17:02	per0830013a
	Perchlorate-O(18)			0.507	ug/L		1	30-AUG-12 17:02	per0830013a

^ When the analyte name is Perchlorate Isotope Ratio the concentration is a unitless value calculated from the ratio of Perchlorate peak area to Perchlorate-101 peak area. The Perchlorate-101 and isotopic ratio results are provided for qualitative purposes only. The results are used to verify the presence and quantitation of Perchlorate.

*Concentration =

$$\text{Instrument Value} \times \frac{\text{Concentrated Extract Volume}}{\text{Aliquot}} \times \frac{1}{\% \text{Solids}}$$

Perchlorate Analysis Data Sheet

Lab Name: GEL Laboratories LLCLab Code: GELInstrument: LCMSMSMethod: SW846 6850 ModifiedMatrix: WATERExtraction Batch ID: 1241187Extraction Type: Filter/DAISample Volume/Weight: 10.0 mLConcentrated Extract Volume: 10.0

Client Sample No.

ICS

Date Received:

GEL Job No (SDG): 12-1508GEL Sample ID: 1202726856Date Filtered: 30-AUG-12Injection Volume (uL): 20

%Solids:

CAS No.	Analyte^	MDL	RL	Conc*	Units	Q	Dilution Factor	Date Analyzed	GEL File ID
14797-73-0	Perchlorate	.05	.2	0.236	ug/L		1	30-AUG-12 17:09	per0830014a
	Perchlorate Isotope Ratio			3.31			1	30-AUG-12 17:09	per0830014a
14797-73-0	Perchlorate-101	.05	.2	0.234	ug/L		1	30-AUG-12 17:09	per0830014a
	Perchlorate-O(18)			0.553	ug/L		1	30-AUG-12 17:09	per0830014a

^ When the analyte name is Perchlorate Isotope Ratio the concentration is a unitless value calculated from the ratio of Perchlorate peak area to Perchlorate-101 peak area. The Perchlorate-101 and isotopic ratio results are provided for qualitative purposes only. The results are used to verify the presence and quantitation of Perchlorate.

*Concentration =

$$\text{Instrument Value} \times \frac{\text{Concentrated Extract Volume}}{\text{Aliquot}} \times \frac{1}{\% \text{Solids}}$$

Perchlorate Analysis Data Sheet

Lab Name: GEL Laboratories LLC

Client Sample No.

CASA-12-21773MSLab Code: GELDate Received: 24-AUG-12Instrument: LCMSMSGEL Job No (SDG): 12-1508Method: SW846 6850 ModifiedGEL Sample ID: 1202726854Matrix: WATERDate Filtered: 30-AUG-12Extraction Batch ID: 1241187Injection Volume (uL): 20Extraction Type: Filter/DAISample Volume/Weight: 10.0 mL

%Solids: .

Concentrated Extract Volume: 10.0

CAS No.	Analyte^	MDL	RL	Conc*	Units	Q	Dilution Factor	Date Analyzed	GEL File ID
14797-73-0	Perchlorate	.05	.2	0.723	ug/L		1	30-AUG-12 18:40	per0830026a
	Perchlorate Isotope Ratio			3.32			1	30-AUG-12 18:40	per0830026a
14797-73-0	Perchlorate-101	.05	.2	0.713	ug/L		1	30-AUG-12 18:40	per0830026a
	Perchlorate-O(18)			0.528	ug/L		1	30-AUG-12 18:40	per0830026a

^ When the analyte name is Perchlorate Isotope Ratio the concentration is a unitless value calculated from the ratio of Perchlorate peak area to Perchlorate-101 peak area. The Perchlorate-101 and isotopic ratio results are provided for qualitative purposes only. The results are used to verify the presence and quantitation of Perchlorate.

*Concentration =

$$\text{Instrument Value} \times \frac{\text{Concentrated Extract Volume}}{\text{Aliquot}} \times \frac{1}{\% \text{Solids}}$$

Perchlorate Analysis Data Sheet

Lab Name: GEL Laboratories LLCLab Code: GELInstrument: LCMSMSMethod: SW846 6850 ModifiedMatrix: WATERExtraction Batch ID: 1241187Extraction Type: Filter/DAISample Volume/Weight: 10.0 mLConcentrated Extract Volume: 10.0

Client Sample No.

CASA-12-21773MSDDate Received: 24-AUG-12GEL Job No (SDG): 12-1508GEL Sample ID: 1202726855Date Filtered: 30-AUG-12Injection Volume (uL): 20

%Solids: .

CAS No.	Analyte^	MDL	RL	Conc*	Units	Q	Dilution Factor	Date Analyzed	GEL File ID
14797-73-0	Perchlorate	.05	.2	0.888	ug/L		1	30-AUG-12 18:47	per0830027a
	Perchlorate Isotope Ratio			3.36			1	30-AUG-12 18:47	per0830027a
14797-73-0	Perchlorate-101	.05	.2	0.867	ug/L		1	30-AUG-12 18:47	per0830027a
	Perchlorate-O(18)			0.532	ug/L		1	30-AUG-12 18:47	per0830027a

^ When the analyte name is Perchlorate Isotope Ratio the concentration is a unitless value calculated from the ratio of Perchlorate peak area to Perchlorate-101 peak area. The Perchlorate-101 and isotopic ratio results are provided for qualitative purposes only. The results are used to verify the presence and quantitation of Perchlorate.

*Concentration =

$$\text{Instrument Value} \times \frac{\text{Concentrated Extract Volume}}{\text{Aliquot}} \times \frac{1}{\% \text{Solids}}$$

Miscellaneous

DATA EXCEPTION REPORT

Mo.Day Yr. 31-AUG-12	Division: Industrial	Quality Criteria: Specifications	Type: Process
Instrument Type: LC-MS/MS	Test / Method: SW846 6850 Modified	Matrix Type: Liquid	Client Code: ESHL
Batch ID: 1241188	Sample Numbers: See Below		
<p>Potentially affected work order(s)(SDG): 309910(12-1509),309911(12-1508),309996(12-1510),310070(12-1511),310180(12-1513),310243(12-1515)</p> <p>Application Issues: Failed Recovery for MSD/PSD</p>			
Specification and Requirements Exception Description:		DER Disposition:	
<p>1. High recoveries for Perchlorate and Perchlorate-101 were observed in MSD (1202726855). The recoveries were 179% and 168%, respectively. The acceptance range is 75-125%.</p>		<p>1. The high recoveries may be the result of the background concentration present in the parent sample, 310180007 (CASA-12-21773). Recoveries in the LCS (1202726853) were acceptable. The data are reported with the appropriate DER.</p>	

Originator's Name:
Charles Wilson 05-SEP-12

Data Validator/Group Leader:
Michael Penny 05-SEP-12

Metals Analysis

Case Narrative

**Metals Fractional Narrative
ARS International (ARSL)
SDG 12-1508**

Sample Analysis

Sample ID	Client ID
309911002	CASA-12-21647
1202724300	Method Blank (MB) ICP
1202724301	Laboratory Control Sample (LCS)
1202724304	309872001(Bay-39L) Serial Dilution (SD)
1202724302	309872001(Bay-39D) Sample Duplicate (DUP)
1202724303	309872001(Bay-39S) Matrix Spike (MS)
1202724292	Method Blank (MB) ICP-MS
1202724293	Laboratory Control Sample (LCS)
1202724296	309872001(Bay-39L) Serial Dilution (SD)
1202724294	309872001(Bay-39D) Sample Duplicate (DUP)
1202724295	309872001(Bay-39S) Matrix Spike (MS)
1202733106	Method Blank (MB) CVAA
1202733107	Laboratory Control Sample (LCS)
1202733112	309910002(CAMO-12-21742L) Serial Dilution (SD)
1202733108	309910002(CAMO-12-21742D) Sample Duplicate (DUP)
1202733109	309910002(CAMO-12-21742S) Matrix Spike (MS)

Method/Analysis Information

Analytical Batch:	1240142, 1240139, 1243838 and 1244727
Prep Batch :	1240141, 1240137 and 1243833
Standard Operating Procedures:	GL-MA-E-013 REV# 21, GL-MA-E-006 REV# 9, GL-MA-E-014 REV# 24, GL-MA-E-010 REV# 25 and GL-GC-E-107 REV# 7
Analytical Method:	SW846 3005/6010B, SW846 3005/6020 DOE-AL, EPA 245.1/245.2 and SM 2340 B
Prep Method :	SW846 3005A and EPA 245.1/245.2 Prep

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

System Configuration

The Hardness as CaCO₃ is calculated from Calcium and Magnesium results.

The Metals analysis-ICP was performed on a PE 7300 Optima radial/axial-viewing inductively coupled plasma atomic emission spectrometer. The instrument is equipped with a Burgener nebulizer, cyclonic spray chamber, and yttrium or scandium internal standard. Operating conditions for the ICP are set at a power level of 1500 watts. The instrument has a peristaltic pump flow rate of 1.4L/min, argon gas flows of 15 L/min and 0.2 L/min for the torch and auxiliary gases, and a flow setting of 0.65L/min for the nebulizer.

The Metals analysis - ICPMS was performed on a Perkin Elmer ELAN 9000 inductively coupled plasma mass spectrometer (ICP-MS). The instrument is equipped with a cross-flow nebulizer, quadrupole mass spectrometer, and dual mode electron multiplier detector. Internal standards of scandium, germanium, indium, tantalum, and/or lutetium were utilized to cover the mass spectrum. Operating conditions are set at 1400W power and combined argon pressures of 360+/- 7 kPa for the plasma and auxiliary gases, and 0.85 L/min carrier gas flow, and an initial lens voltage of 5.2.

The Metals analysis-Mercury was performed on a Perkin-Elmer Flow Injection Mercury System (FIMS-100) automated mercury analyzer. The instrument consists of a cold vapor atomic absorption spectrometer set to detect mercury at a wavelength of 253.7 nm. Sample introduction through the flow injection system is performed via a peristaltic pump at 9 mL/min and nitrogen carrier gas rate of 80 mL/min.

Calibration Information

Instrument Calibration

All initial calibration requirements have been met for this sample delivery group (SDG).

CRDL Requirements

All CRDL standard(s) met the referenced advisory control limits with the exception of potassium that recovered outside of the advisory control limit.

ICSA/ICSAB Statement

All interference check samples (ICSA and ICSAB) associated with this SDG met the established acceptance criteria.

Continuing Calibration Blank (CCB) Requirements

All continuing calibration blanks (CCB) bracketing this batch met the established acceptance

criteria.

Continuing Calibration Verification (CCV) Requirements

All continuing calibration verifications (CCV) bracketing this SDG met the acceptance criteria.

Quality Control (QC) Information

Method Blank (MB) Statement

The MBs analyzed with this SDG met the acceptance criteria.

Laboratory Control Sample (LCS) Recovery

The LCS spike recoveries met the acceptance limits.

Quality Control (QC) Sample Statement

The following samples were selected as the quality control (QC) samples for this SDG: 309872001 (Bay-39)-ICP and ICP-MS and 309910002 (CAMO-12-21742)-CVAA.

Matrix Spike (MS) Recovery Statement

The percent recoveries (%R) obtained from the MS analyses are evaluated when the sample concentration is less than four times (4X) the spike concentration added. All applicable elements met the acceptance criteria.

Duplicate Relative Percent Difference (RPD) Statement

The relative percent difference (RPD) obtained from the designated sample duplicate (DUP) is evaluated based on acceptance criteria of 20% when the sample is >5X the contract required detection limit (RL). In cases where either the sample or duplicate value is less than 5X the contract required detection limit (RL), a control of +/-RL is used to evaluate the DUP results. All applicable analytes met these requirements with the exception of iron.

Serial Dilution % Difference Statement

The serial dilution is used to assess matrix suppression or enhancement. Raw element concentrations that are 25X the IDL/MDL for CVAA, 50X the IDL/MDL for ICP, and 100X the IDL/MDL for ICP-MS analyses are applicable for serial dilution assessment. All applicable analytes met the acceptance criteria of less than 10% difference (%D).

Technical Information

Holding Time Specifications

GEL assigns holding times based on the associated methodology, which assigns the date and time from sample collection of sample receipt. Those holding times expressed in hours are calculated in the AlphaLIMS system. Those holding times expressed as days expire at midnight on the day of expiration. All samples in this SDG met the specified holding time.

Preparation/Analytical Method Verification

All procedures were performed as stated in the SOP.

Sample Dilutions

Dilutions are performed to minimize matrix interferences resulting from elevated mineral element concentrations present in solid samples and/or to bring over range target analyte concentrations into the linear calibration range of the instrument. The samples in this SDG did not require dilutions.

Preparation Information

The samples in this SDG were prepared exactly according to the cited SOP.

Miscellaneous Information

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. An electronic signature page inserted after the case narrative will include the data validator's signature and title. The signature page also includes the data qualifiers used in the fractional package. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Data Exception (DER) Documentation

Data exception reports are generated to document any procedural anomalies that may deviate from referenced SOP or contractual documents. The following DER was generated for this SDG: DER ID 1118029. A copy is included in the Miscellaneous Data section of this package.

Additional Comments

Additional comments were not required for this SDG.

Total Hardness by Calculation is determined using the results of Total Calcium (Ca) and Total Magnesium (Mg) determined by ICP or ICP-MS.

$$\text{Hardness} = 2.497 (\text{Ca}) + 4.118 (\text{Mg})$$

Please refer to the Total Ca and Total Mg data to validate results appearing on the Hardness Summary sheet. Both results are in the Inorganic/metals section of the package. There is no Batch QC for calculated results, and thus no QC Summary for the Hardness by Calculation Batch. The MDLs and PQLs are calculated using the higher of the two calculated values of Ca or Mg.

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case

Sample Data Summary

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Qualifier Definition Report for

ARSL001 ARS International (63641-10)

Client SDG: 12-1508 GEL Work Order: 309911

The Qualifiers in this report are defined as follows:

- * A quality control analyte recovery is outside of specified acceptance criteria
- ** Analyte is a surrogate compound
- J Value is estimated
- U Analyte was analyzed for, but not detected above the MDL, MDA, or LOD.

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless qualified on the Certificate of Analysis.

The designation ND, if present, appears in the result column when the analyte concentration is not detected above the detection limit.

This data report has been prepared and reviewed in accordance with GEL Laboratories LLC standard operating procedures. Please direct any questions to your Project Manager, Valerie Davis.

Reviewed by



09.12.12

METALS
-1-
INORGANICS ANALYSIS DATA PACKAGE

SDG No: 12-1508

METHOD TYPE: EPA

SAMPLE ID: 309911002

CLIENT ID: CASA-12-21647

CONTRACT: ESHL00210

MATRIX:W

DATE RECEIVED 21-AUG-12

LEVEL: Low **%SOLIDS:**

<u>CAS No</u>	<u>Analyte</u>	<u>Result</u>	<u>Units</u>	<u>C</u>	<u>Qual</u>	<u>M*</u>	<u>MDL</u>	<u>DF</u>	<u>Inst ID</u>	<u>Analytical Run</u>
7439-97-6	Mercury	0.067	ug/L	U		AV	0.067	1	MER536	090712W1-5
7429-90-5	Aluminum	68	ug/L	U		P	68	1	OPTIMA5	090512B-1
7440-36-0	Antimony	1	ug/L	U		MS	1	1	ICPMS5	120907-2
7440-38-2	Arsenic	2.56	ug/L	J		MS	1.7	1	ICPMS5	120907-2
7440-39-3	Barium	39.8	ug/L			P	1	1	OPTIMA5	090512B-1
7440-41-7	Beryllium	1	ug/L	U		P	1	1	OPTIMA5	090512B-1
7440-42-8	Boron	28.9	ug/L	J		P	15	1	OPTIMA5	090512B-1
7440-43-9	Cadmium	0.11	ug/L	U		MS	0.11	1	ICPMS5	120907-2
7440-70-2	Calcium	22700	ug/L			P	50	1	OPTIMA5	090512B-1
7440-47-3	Chromium	20.3	ug/L			MS	2	1	ICPMS5	120907-2
7440-48-4	Cobalt	1	ug/L	U		P	1	1	OPTIMA5	090512B-1
7440-50-8	Copper	3	ug/L	U		P	3	1	OPTIMA5	090512B-1
7439-89-6	Iron	30	ug/L	U		P	30	1	OPTIMA5	090512B-1
7439-92-1	Lead	0.5	ug/L	U		MS	0.5	1	ICPMS5	120911-4
7439-95-4	Magnesium	6080	ug/L			P	110	1	OPTIMA5	090512B-1
7439-96-5	Manganese	2	ug/L	U		P	2	1	OPTIMA5	090512B-1
7439-98-7	Molybdenum	1.46	ug/L			MS	0.165	1	ICPMS5	120907-2
7440-02-0	Nickel	0.617	ug/L	J		MS	0.5	1	ICPMS5	120910-3
7440-09-7	Potassium	1540	ug/L			P	50	1	OPTIMA5	090512B-1
7782-49-2	Selenium	1.76	ug/L	J		MS	1.5	1	ICPMS5	120907-2
7631-86-9	Silica	70800	ug/L			P	53	1	OPTIMA5	090512B-1
7440-22-4	Silver	0.2	ug/L	U		MS	0.2	1	ICPMS5	120907-2
7440-23-5	Sodium	12900	ug/L			P	100	1	OPTIMA5	090512B-1
7440-24-6	Strontium	96.6	ug/L			P	1	1	OPTIMA5	090512B-1
7440-28-0	Thallium	0.45	ug/L	U		MS	0.45	1	ICPMS5	120907-2
7440-31-5	Tin	2.5	ug/L	U		P	2.5	1	OPTIMA5	090512B-1

METALS
-1-
INORGANICS ANALYSIS DATA PACKAGE

SDG No: 12-1508

METHOD TYPE: EPA

SAMPLE ID: 309911002

CLIENT ID: CASA-12-21647

CONTRACT: ESHL00210

MATRIX:W

DATE RECEIVED 21-AUG-12

LEVEL: Low **%SOLIDS:**

<u>CAS No</u>	<u>Analyte</u>	<u>Result</u>	<u>Units</u>	<u>C</u>	<u>Qual</u>	<u>M*</u>	<u>MDL</u>	<u>DF</u>	<u>Inst ID</u>	<u>Analytical Run</u>
7440-61-1	Uranium	0.714	ug/L			MS	0.067	1	ICPMS5	120907-2
7440-62-2	Vanadium	6.54	ug/L			P	1	1	OPTIMA5	090512B-1
7440-66-6	Zinc	9.79	ug/L	J		P	3.3	1	OPTIMA5	090512B-1
	Hardness as CaCO3	81.8	mg/L				0.453	1	CALC001	

***Analytical Methods:**

- MS SW846 3005/6020 DOE-AL**
- P SW846 3005/6010B**
- AV EPA 245.1/245.2**
- SM 2340 B**

Quality Control Summary

METALS
-3b-
PREPARATION BLANK SUMMARY

SDG NO. 12-1508
Contract: ESHL00210
Matrix: W

<u>Sample ID</u>	<u>Analyte</u>	<u>Result</u>	<u>Units</u>	<u>Acceptance Window</u>	<u>Conc Qual</u>	<u>M*</u>	<u>MDL</u>	<u>RDL</u>
1202724292								
	Antimony	1	ug/L	+/-3	U	MS	1	3
	Arsenic	1.7	ug/L	+/-5	U	MS	1.7	5
	Cadmium	0.11	ug/L	+/-1	U	MS	0.11	1
	Chromium	2	ug/L	+/-10	U	MS	2	10
	Lead	0.5	ug/L	+/-2	U	MS	0.5	2
	Molybdenum	0.165	ug/L	+/-0.5	U	MS	0.165	0.5
	Nickel	0.5	ug/L	+/-2	U	MS	0.5	2
	Selenium	1.5	ug/L	+/-5	U	MS	1.5	5
	Silver	0.2	ug/L	+/-1	U	MS	0.2	1
	Thallium	0.45	ug/L	+/-2	U	MS	0.45	2
	Uranium	0.067	ug/L	+/-0.2	U	MS	0.067	0.2
1202724300								
	Aluminum	68	ug/L	+/-200	U	P	68	200
	Barium	1	ug/L	+/-5	U	P	1	5
	Beryllium	1	ug/L	+/-5	U	P	1	5
	Boron	15	ug/L	+/-50	U	P	15	50
	Calcium	50	ug/L	+/-200	U	P	50	200
	Cobalt	1	ug/L	+/-5	U	P	1	5
	Copper	3	ug/L	+/-10	U	P	3	10
	Iron	30	ug/L	+/-100	U	P	30	100
	Magnesium	110	ug/L	+/-300	U	P	110	300
	Manganese	2	ug/L	+/-10	U	P	2	10
	Potassium	-153	ug/L	+/-150		P	50	150
	Silica	53	ug/L	+/-213	U	P	53	213
	Sodium	100	ug/L	+/-300	U	P	100	300
	Strontium	1	ug/L	+/-5	U	P	1	5
	Tin	2.5	ug/L	+/-10	U	P	2.5	10
	Vanadium	1	ug/L	+/-5	U	P	1	5
	Zinc	3.3	ug/L	+/-10	U	P	3.3	10
1202733106								
	Mercury	-0.085	ug/L	+/-0.2	J	AV	0.067	0.2

*Analytical Methods:

MS SW846 3005/6020 DOE-AL
P SW846 3005/6010B
AV EPA 245.1/245.2

METALS

-5a-

Matrix Spike Summary

SDG NO. 12-1508 Client ID: Bay-39S
 Contract: LANL00110 Level: Low
 Matrix: WATER % Solids:
 Sample ID: 309872001 Spike ID: 1202724295

Analyte	Units	Acceptance Limit	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M*
Antimony	ug/L	75-125	202		1.8	J	200	100		MS
Arsenic	ug/L	75-125	83		1.7	U	80	102		MS
Cadmium	ug/L		122		110		10	120	N/A	MS
Chromium	ug/L	75-125	74.6		24.5		50	100		MS
Lead	ug/L	75-125	116		77.5		40	97.3		MS
Molybdenum	ug/L	75-125	51.7		0.165	U	50	103		MS
Nickel	ug/L	75-125	61		6.84		50	108		MS
Selenium	ug/L	75-125	25.4		6.1		20	96.6		MS
Silver	ug/L	75-125	52.7		0.2	U	50	105		MS
Thallium	ug/L	75-125	102		0.45	U	100	102		MS
Uranium	ug/L	75-125	51.6		0.071	J	50	103		MS

*Analytical Methods:

MS SW846 3005/6020 DOE-AL

METALS

-5a-

Matrix Spike Summary

SDG NO. 12-1508 Client ID: Bay-39S
 Contract: LANL00110 Level: Low
 Matrix: WATER % Solids:
 Sample ID: 309872001 Spike ID: 1202724303

Analyte	Units	Acceptance Limit	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M*
Aluminum	ug/L	75-125	5150		68	U	5000	102		P
Barium	ug/L	75-125	755		258		500	99.4		P
Beryllium	ug/L	75-125	488		1	U	500	97.6		P
Boron	ug/L	75-125	660		157		500	101		P
Calcium	ug/L	75-125	8530		3630		5000	98		P
Cobalt	ug/L	75-125	511		1.86	J	500	102		P
Copper	ug/L	75-125	513		4.85	J	500	102		P
Iron	ug/L	75-125	5300		762		5000	90.8		P
Magnesium	ug/L	75-125	6350		1540		5000	96.2		P
Manganese	ug/L	75-125	586		103		500	96.7		P
Potassium	ug/L	75-125	6580		1440		5000	103		P
Silica	ug/L	75-125	11600		1960		10700	90.1		P
Sodium	ug/L	75-125	5620		513		5000	102		P
Strontium	ug/L	75-125	527		19.3		500	102		P
Tin	ug/L	75-125	494		2.5	U	500	98.8		P
Vanadium	ug/L	75-125	505		1	U	500	101		P
Zinc	ug/L	75-125	916		447		500	93.6		P

*Analytical Methods:

P SW846 3005/6010B

METALS

-5a-

Matrix Spike Summary

SDG NO. 12-1508 Client ID: CAMO-12-21742S

Contract: ESHL00210 Level: Low

Matrix: WATER % Solids:

Sample ID: 309910002 Spike ID: 1202733109

<u>Analyte</u>	<u>Units</u>	<u>Acceptance Limit</u>	<u>Spiked Result</u>	<u>C</u>	<u>Sample Result</u>	<u>C</u>	<u>Spike Added</u>	<u>% Recovery</u>	<u>Qual</u>	<u>M*</u>
Mercury	ug/L	75-125	2.01		0.067	U	2	100		AV

*Analytical Methods:

AV EPA 245.1/245.2

Metals
-6-
Duplicate Sample Summary

SDG No.: 12-1508

Lab Code: GEL

Contract: ESHL00210

Client ID: Bay-39D

Matrix: LIQUID

Level: Low

Sample ID: 309872001

Duplicate ID: 1202724294

Percent Solids for Dup: N/A

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M*
Antimony	ug/L	+/-3	1.8 J		1.6 J		12		MS
Arsenic	ug/L		1.7 U		1.7 U				MS
Cadmium	ug/L	+/-20%	110		103		6.99		MS
Chromium	ug/L	+/-10	24.5		23.1		5.95		MS
Lead	ug/L	+/-20%	77.5		74.3		4.26		MS
Molybdenum	ug/L		0.165 U		0.165 U				MS
Nickel	ug/L	+/-2	6.84		6.24		9.15		MS
Selenium	ug/L	+/-5	6.1		5.64		7.72		MS
Silver	ug/L		0.2 U		0.2 U				MS
Thallium	ug/L		0.45 U		0.45 U				MS
Uranium	ug/L		0.071 J		0.067 U		200		MS

*Analytical Methods:

MS SW846 3005/6020 DOE-AL

Metals
-6-
Duplicate Sample Summary

SDG No.: 12-1508

Lab Code: GEL

Contract: ESHL00210

Client ID: Bay-39D

Matrix: LIQUID

Level: Low

Sample ID: 309872001

Duplicate ID: 1202724302

Percent Solids for Dup: N/A

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M*
Aluminum	ug/L		68 U		68 U				P
Barium	ug/L	+/-20%	258		265		2.84		P
Beryllium	ug/L		1 U		1 U				P
Boron	ug/L	+/-50	157		161		2.42		P
Calcium	ug/L	+/-20%	3630		3720		2.42		P
Cobalt	ug/L	+/-5	1.86 J		1.65 J		12.1		P
Copper	ug/L	+/-10	4.85 J		4.79 J		1.22		P
Iron	ug/L	+/-20%	762		582		26.8	*	P
Magnesium	ug/L	+/-20%	1540		1560		1.58		P
Manganese	ug/L	+/-20%	103		105		1.96		P
Potassium	ug/L	+/-20%	1440		1550		7.26		P
Silica	ug/L	+/-20%	1960		1940		.961		P
Sodium	ug/L	+/-300	513		535		4.15		P
Strontium	ug/L	+/-5	19.3		19.7		1.75		P
Tin	ug/L		2.5 U		2.5 U				P
Vanadium	ug/L		1 U		1 U				P
Zinc	ug/L	+/-20%	447		457		2.12		P

*Analytical Methods:

P SW846 3005/6010B

Metals
-6-
Duplicate Sample Summary

SDG No.: 12-1508

Lab Code: GEL

Contract: ESHL00210

Client ID: CAMO-12-21742D

Matrix: LIQUID

Level: Low

Sample ID: 309910002

Duplicate ID: 1202733108

Percent Solids for Dup: N/A

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M*
Mercury	ug/L		0.067	U	0.067	U			AV

*Analytical Methods:

AV EPA 245.1/245.2

METALS

-7-

Laboratory Control Sample Summary

SDG NO. 12-1508

Contract: ESHL00210

Aqueous LCS Source:O2si

Solid LCS Source:

<u>Sample ID</u>	<u>Analyte</u>	<u>Units</u>	<u>True Value</u>	<u>Result</u>	<u>C</u>	<u>% Recovery</u>	<u>Acceptance Limit</u>	<u>M*</u>
1202724293								
	Antimony	ug/L	50	49.6		99.2	80-120	MS
	Arsenic	ug/L	50	50.9		102	80-120	MS
	Cadmium	ug/L	50	50.2		100	80-120	MS
	Chromium	ug/L	50	53.7		107	80-120	MS
	Lead	ug/L	50	52.1		104	80-120	MS
	Molybdenum	ug/L	50	48.8		97.7	80-120	MS
	Nickel	ug/L	50	54.3		109	80-120	MS
	Selenium	ug/L	50	50.2		100	80-120	MS
	Silver	ug/L	50	52		104	80-120	MS
	Thallium	ug/L	50	51.1		102	80-120	MS
	Uranium	ug/L	50	52		104	80-120	MS

*Analytical Methods:

MS SW846 3005/6020 DOE-AL

METALS

-7-

Laboratory Control Sample Summary

SDG NO. 12-1508

Contract: ESHL00210

Aqueous LCS Source:OS2I

Solid LCS Source:

<u>Sample ID</u>	<u>Analyte</u>	<u>Units</u>	<u>True Value</u>	<u>Result</u>	<u>C</u>	<u>% Recovery</u>	<u>Acceptance Limit</u>	<u>M*</u>
1202724301	Aluminum	ug/L	5000	5290		106	80-120	P
	Barium	ug/L	500	520		104	80-120	P
	Beryllium	ug/L	500	515		103	80-120	P
	Boron	ug/L	500	531		106	80-120	P
	Calcium	ug/L	5000	5090		102	80-120	P
	Cobalt	ug/L	500	528		106	80-120	P
	Copper	ug/L	500	526		105	80-120	P
	Iron	ug/L	5000	5010		100	80-120	P
	Magnesium	ug/L	5000	5070		101	80-120	P
	Manganese	ug/L	500	512		102	80-120	P
	Potassium	ug/L	5000	5300		106	80-120	P
	Silica	ug/L	10700	10200		95.6	80-120	P
	Sodium	ug/L	5000	5350		107	80-120	P
	Strontium	ug/L	500	531		106	80-120	P
	Tin	ug/L	500	521		104	80-120	P
	Vanadium	ug/L	500	528		106	80-120	P
	Zinc	ug/L	500	513		103	80-120	P

*Analytical Methods:

P SW846 3005/6010B

METALS

-7-

Laboratory Control Sample Summary

SDG NO. 12-1508

Contract: ESHL00210

Aqueous LCS Source:GEL

Solid LCS Source:

<u>Sample ID</u>	<u>Analyte</u>	<u>Units</u>	<u>True Value</u>	<u>Result</u>	<u>C</u>	<u>% Recovery</u>	<u>Acceptance Limit</u>	<u>M*</u>
1202733107	Mercury	ug/L	2	2.07		103	85-115	AV

*Analytical Methods:

AV EPA 245.1/245.2

METALS

-9-

Serial Dilution Sample Summary

SDG NO. 12-1508 Client ID: Bay-39L

Contract: ESHL00210

Matrix: LIQUID Level: Low

Sample ID: 309872001 Serial Dilution ID: 1202724296

<u>Analyte</u>	<u>Initial Value</u> ug/L	<u>C</u>	<u>Serial Value</u> ug/L	<u>C</u>	<u>% Difference</u>	<u>Qual</u>	<u>Acceptance Limit</u>	<u>M*</u>
Antimony	1.8	J	5	U	100			MS
Arsenic	1.7	U	8.5	U				MS
Cadmium	110		106		3.43		10	MS
Chromium	24.5		26.3	J	7			MS
Lead	77.5		76.3		1.6		10	MS
Molybdenum	.165	U	.825	U				MS
Nickel	6.84		6.34	J	7.33			MS
Selenium	6.1		7.5	U	100			MS
Silver	.2	U	1	U				MS
Thallium	.45	U	2.25	U				MS
Uranium	.071	J	.335	U	100			MS

*Analytical Methods:

MS SW846 3005/6020 DOE-AL

METALS

-9-

Serial Dilution Sample Summary

SDG NO. 12-1508

Client ID: Bay-39L

Contract: ESHL00210

Matrix: LIQUID

Level: Low

Sample ID: 309872001

Serial Dilution ID: 1202724304

<u>Analyte</u>	<u>Initial Value</u> ug/L	<u>C</u>	<u>Serial Value</u> ug/L	<u>C</u>	<u>% Difference</u>	<u>Qual</u>	<u>Acceptance Limit</u>	<u>M*</u>
Aluminum	68	U	340	U				P
Barium	258		264		2.3		10	P
Beryllium	1	U	5	U				P
Boron	157		168	J	6.54			P
Calcium	3630		3740		3.16		10	P
Cobalt	1.86	J	5	U	100			P
Copper	4.85	J	15	U	100			P
Iron	762		566		25.8			P
Magnesium	1540		1650		7.38			P
Manganese	103		105		2.78		10	P
Potassium	1440		932		35.4			P
Silica	1960		1890		3.25			P
Sodium	513		500	U	100			P
Strontium	19.3		19	J	1.65			P
Tin	2.5	U	12.5	U				P
Vanadium	1	U	5	U				P
Zinc	447		464		3.72		10	P

*Analytical Methods:

P SW846 3005/6010B

METALS

-9-

Serial Dilution Sample Summary

SDG NO. 12-1508 Client ID: CAMO-12-21742L

Contract: ESHL00210

Matrix: LIQUID Level: Low

Sample ID: 309910002 Serial Dilution ID: 1202733112

<u>Analyte</u>	<u>Initial Value</u> ug/L	<u>C</u>	<u>Serial Value</u> ug/L	<u>C</u>	<u>% Difference</u>	<u>Qual</u>	<u>Acceptance Limit</u>	<u>M*</u>
Mercury	.067	U	.335	U				AV

*Analytical Methods:

AV EPA 245.1/245.2

Miscellaneous

DATA EXCEPTION REPORT

Mo.Day Yr. 06-SEP-12	Division: Industrial	Quality Criteria: Specifications	Type: Process
Instrument Type: ICP	Test / Method: SW846 3005/6010B	Matrix Type: Liquid	Client Code: ESHL, LANL
Batch ID: 1240142	Sample Numbers: See Below		
Potentially affected work order(s)(SDG): 309872,309910(12-1509),309911(12-1508)			
Application Issues: Failed RPD for DUP			
Specification and Requirements Exception Description:		DER Disposition:	
<p>1. Failed RPD for DUP: QC 1202724302DUP</p>		<p>1. The sample and sample duplicate % RPD failed outside the control limits for iron due to possible sample non-homogeneity and/or matrix interference. Per GEL's accredited methods and SOPs, a corrective action is not required and the data is qualified and reported.</p>	

Originator's Name:

Jerry Wigfall 06-SEP-12

Data Validator/Group Leader:

Jamie Johnson 07-SEP-12

General Chem Analysis

Case Narrative

**General Chemistry Narrative
ARS International (ARSL)
SDG 12-1508**

Method/Analysis Information

Product: Carbon, Total Organic

Analytical Batch: 1241231

Method: SW 9060 Total Organic Carbon

Sample Analysis

The following samples were analyzed using the analytical protocol as established in SW846 9060:

Sample ID	Client ID
309911001	CASA-12-21643
1202726977	Method Blank (MB)
1202726978	309911001(CASA-12-21643) Sample Duplicate (DUP)
1202726979	310243006(CAMO-12-21779) Sample Duplicate (DUP)
1202726980	309911001(CASA-12-21643) Post Spike (PS)
1202726981	310243006(CAMO-12-21779) Post Spike (PS)
1202726982	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-093 REV# 10.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Carbon analysis was performed on a O-I Analytical Model 1010 Total Organic Carbon Analyzer.

Initial Calibration

All initial calibration requirements have been met for this SDG.

Continuing Calibration Blanks

All continuing calibration blanks (CCBs) associated with reported data from this batch were within acceptance limits.

Calibration Verification Information (CCV)

All continuing calibration verification standards (CCVs) associated with reported data from this batch were within acceptance limits.

Quality Control (QC) Information**Method Blank (MB) Statement**

The MB analyzed with this SDG met the acceptance criteria.

Laboratory Control Sample (LCS) Recovery

The LCS spike recovery met the acceptance limits.

Quality Control (QC) Designation

The following samples were selected for QC analysis: 309911001 (CASA-12-21643) and 310243006 (CAMO-12-21779).

Matrix Spike (MS)/Post Spike (PS) Recovery Statement

The MS/PS recoveries for this sample set were within the required acceptance limits.

Duplicate Relative Percent Difference (RPD) Statement

The RPD between the sample and its duplicate met the acceptance limits.

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

All samples in this SDG met the specified holding time.

Sample Preservation/Integrity

All the samples from this sample group met the preservation and integrity requirements of the method.

Sample Dilutions

The samples in this SDG did not require dilutions.

Sample Re-analysis

The following samples were re-analyzed due to instrument failure: 1202726979 (CAMO-12-21779) and 1202726981 (CAMO-12-21779).

Miscellaneous Information**Data Exception (DER) Documentation**

Data exception reports (DERs) are generated to document procedural anomalies that may deviate from referenced SOP or contractual documents. A data exception report (DER) was not generated for this SDG.

Additional Comments

A 15 mg/L Total Inorganic Carbon check standard is analyzed with each analytical run to prove that the instrument is effectively sparging away the inorganic carbon.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Method/Analysis Information

Product: Specific Conductivity
Analytical Batch: 1242995 **Method:** EPA120.1 Specific Conductivity

Sample Analysis

The following samples were analyzed using the analytical protocol as established in EPA 120.1:

Sample ID	Client ID
309911002	CASA-12-21647
1202731308	309910002(CAMO-12-21742) Sample Duplicate (DUP)
1202731309	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-009 REV# 11.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Conductivity analysis was performed on a Orion 160 Conductivity Meter.

Initial Calibration

All initial calibration requirements have been met for this SDG.

Calibration Verification Information

All associated calibration verification standard(s) (ICV or CCV) met the acceptance criteria.

Quality Control (QC) Information

Laboratory Control Sample (LCS) Recovery

The LCS spike recovery met the acceptance limits.

Quality Control (QC) Designation

The following sample was selected for QC analysis: 309910002 (CAMO-12-21742).

Duplicate Relative Percent Difference (RPD) Statement

The RPD between the sample and its duplicate met the acceptance limits.

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

All samples in this SDG met the specified holding time.

Sample Dilutions

The samples in this SDG did not require dilutions.

Sample Re-analysis

The samples in this SDG did not require re-analysis.

Miscellaneous Information**Data Exception (DER) Documentation**

Data exception reports (DERs) are generated to document procedural anomalies that may deviate from referenced SOP or contractual documents. A data exception report (DER) was not generated for this SDG.

Additional Comments

Additional comments were not required for this SDG.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Method/Analysis Information

Product: pH
Analytical Batch: 1241830 **Method:** EPA 150.1 pH

Sample Analysis

The following samples were analyzed using the analytical protocol as established in EPA 150.1:

Sample ID	Client ID
309911002	CASA-12-21647
1202728424	309907006(WTRO-12-23145) Sample Duplicate (DUP)
1202728425	309910002(CAMO-12-21742) Sample Duplicate (DUP)
1202728426	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-008 REV# 21.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Electrode analysis was performed on a PerpHect pH Meter Orion 370.

Initial Calibration

All initial calibration requirements have been met for this SDG.

Calibration Verification Information (CCV)

All continuing calibration verification standards (CCVs) associated with reported data from this batch were within acceptance limits.

Quality Control (QC) Information

Laboratory Control Sample (LCS) Recovery

The LCS spike recovery met the acceptance limits.

Quality Control (QC) Designation

The following samples were selected for QC analysis: 309907006 (WTRO-12-23145) and 309910002 (CAMO-12-21742).

Duplicate Relative Percent Difference (RPD) Statement

The RPD between the sample and its duplicate met the acceptance limits.

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

The following sample from this sample group was received by the lab outside of the method specified holding time: 309911002 (CASA-12-21647).

Sample Re-analysis

The samples in this SDG did not require re-analysis.

Miscellaneous Information**Data Exception (DER) Documentation**

The following DER was generated for this SDG: 1117658 309911002 (CASA-12-21647).

Additional Comments

Additional comments were not required for this SDG.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Method/Analysis Information

Product: Ion Chromatography
Analytical Batch: 1241176 **Method:** EPA 300.0 Anions Liquid 28 day

Sample Analysis

The following samples were analyzed using the analytical protocol as established in EPA 300.0:

Sample ID	Client ID
309911002	CASA-12-21647
1202726829	Method Blank (MB)
1202726830	310180003(CASA-12-21772) Sample Duplicate (DUP)
1202726831	310180003(CASA-12-21772) Post Spike (PS)
1202726832	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-086 REV# 21.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Ion Chromatography analysis was performed on a Dionex ICS-3000 Ion Chromatograph.

Initial Calibration

All initial calibration requirements have been met for this SDG.

Continuing Calibration Blanks

All continuing calibration blanks (CCBs) associated with reported data from this batch were within acceptance limits.

Calibration Verification Information (CCV)

All continuing calibration verification standards (CCVs) associated with reported data from this batch were within acceptance limits.

Y Intercept Rule

The absolute value of the intercept is less than 3 times the MDL.

Quality Control (QC) Information**Method Blank (MB) Statement**

The MB analyzed with this SDG met the acceptance criteria.

Laboratory Control Sample (LCS) Recovery

The LCS spike recovery met the acceptance limits.

Quality Control (QC) Designation

The following sample was selected for QC analysis: 310180003 (CASA-12-21772).

Matrix Spike (MS)/Post Spike (PS) Recovery Statement

The MS/PS recovery for this sample set was within the required acceptance limits.

Duplicate Relative Percent Difference (RPD) Statement

The RPD between the sample and its duplicate met the acceptance limits.

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

All samples in this SDG met the specified holding time.

Sample Dilutions

The samples in this SDG did not require dilutions.

Sample Re-analysis

The samples in this SDG did not require re-analysis.

Miscellaneous Information**Data Exception (DER) Documentation**

Data exception reports (DERs) are generated to document procedural anomalies that may deviate from referenced SOP or contractual documents. A data exception report (DER) was not generated for this SDG.

Manual Integrations

The following samples from this sample group had to be manually integrated due to errors in the instrument software peak integration: 1202726830 (CASA-12-21772), 1202726831 (CASA-12-21772) and 309911002 (CASA-12-21647).

Additional Comments

Additional comments were not required for this SDG.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an

effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Method/Analysis Information

Product: Ammonia Nitrogen
Analytical Batch: 1240583 **Method:** EPA 350.1 Nitrogen and Ammonia L
Prep Batch : 1240582 **Method:** EEPA 350.2 Prep

Sample Analysis

The following samples were analyzed using the analytical protocol as established in EPA 350.1:

Sample ID	Client ID
309911002	CASA-12-21647
1202725417	Method Blank (MB)
1202725418	309548008(CAMO-12-21796) Sample Duplicate (DUP)
1202725420	309548008(CAMO-12-21796) Matrix Spike (MS)
1202725422	309548008(CAMO-12-21796) Matrix Spike Duplicate (MSD)
1202725424	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-106 REV# 8.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Nutrient analysis was performed on a Lachat QuickChem FIA+ 8000 Series.

Continuing Calibration Blanks

All continuing calibration blanks (CCBs) associated with reported data from this batch were within acceptance limits.

Calibration Verification Information (CCV)

All continuing calibration verification standards (CCVs) associated with reported data from this batch were within acceptance limits.

Y Intercept Rule

The absolute value of the intercept is less than 3 times the MDL.

Quality Control (QC) Information

Method Blank (MB) Statement

The MB analyzed with this SDG met the acceptance criteria.

Laboratory Control Sample (LCS) Recovery

The LCS spike recovery met the acceptance limits.

Quality Control (QC) Designation

The following sample was selected for QC analysis: 309548008 (CAMO-12-21796).

Matrix Spike (MS)/Post Spike (PS) Recovery Statement

The MS/PS recovery for this sample set was within the required acceptance limits.

Matrix Spike Duplicate (MSD) Recovery Statement

The MSD recovery for this sample set was within the required acceptance limits.

MS/MSD Relative Percent Difference (RPD) Statement

The RPD between the spike and spike duplicate met the acceptance limits.

Duplicate Relative Percent Difference (RPD) Statement

One or more of the values for the sample and/or duplicate are less than 5 times the Practical Quantitation Limit (PQL), and the difference is within one PQL value; therefore, the RPD is not applicable. 1202725418 (CAMO-12-21796).

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

All samples in this SDG met the specified holding time.

Sample Preservation/Integrity

All the samples from this sample group met the preservation and integrity requirements of the method.

Sample Dilutions

The samples in this SDG did not require dilutions.

Sample Re-analysis

The samples in this SDG did not require re-analysis.

Miscellaneous Information

Data Exception (DER) Documentation

Data exception reports (DERs) are generated to document procedural anomalies that may deviate from referenced SOP or contractual documents. A data exception report (DER) was not generated for this SDG.

Additional Comments

Additional comments were not required for this SDG.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Method/Analysis Information

Product: Total Kjeldahl Nitrogen
Analytical Batch: 1240586 **Method:** Nitrogen and Total Kjeldahl (TKN)
Prep Batch : 1240585 **Method:** EEPA 351.2 Prep

Sample Analysis

The following samples were analyzed using the analytical protocol as established in EPA 351.2:

Sample ID	Client ID
309911001	CASA-12-21643
1202725433	Method Blank (MB)
1202725434	309791001(SWWS46-12-22930) Sample Duplicate (DUP)
1202725435	309791001(SWWS46-12-22930) Matrix Spike (MS)
1202725436	309791001(SWWS46-12-22930) Matrix Spike Duplicate (MSD)
1202725437	Laboratory Control Sample (LCS)
1202730363	309910001(CAMO-12-21734) Sample Duplicate (DUP)
1202730364	309910001(CAMO-12-21734) Matrix Spike (MS)
1202730365	309910001(CAMO-12-21734) Matrix Spike Duplicate (MSD)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-104 REV# 12.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Nutrient analysis was performed on a Lachat QuickChem FIA+ 8000 Series.

Continuing Calibration Blanks

All continuing calibration blanks (CCBs) associated with reported data from this batch were within acceptance limits.

Calibration Verification Information (CCV)

All continuing calibration verification standards (CCVs) associated with reported data from this batch were within

acceptance limits.

Y Intercept Rule

The absolute value of the intercept is less than 3 times the MDL.

Quality Control (QC) Information

Method Blank (MB) Statement

The MB analyzed with this SDG met the acceptance criteria.

Laboratory Control Sample (LCS) Recovery

The LCS spike recovery met the acceptance limits.

Quality Control (QC) Designation

The following samples were selected for QC analysis: 309791001 (SWWS46-12-22930) and 309910001 (CAMO-12-21734).

Matrix Spike (MS)/Post Spike (PS) Recovery Statement

The spike recovery falls outside of the established acceptance limits. Since both the spike duplicate recovery and the RPD between the spike and spike duplicate fall within acceptance limits, the data is reported. 1202730364 (CAMO-12-21734).

Matrix Spike Duplicate (MSD) Recovery Statement

The MSD recoveries for this sample set were within the required acceptance limits.

MS/MSD Relative Percent Difference (RPD) Statement

The RPDs between the spike and spike duplicate met the acceptance limits.

Duplicate Relative Percent Difference (RPD) Statement

The values for the sample and duplicate are less than the Practical Quantitation Limit (PQL); therefore, the RPD is not applicable. 1202730363 (CAMO-12-21734).

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

All samples in this SDG met the specified holding time.

Sample Preservation/Integrity

All the samples from this sample group met the preservation and integrity requirements of the method.

Sample Dilutions

The samples in this SDG did not require dilutions.

Sample Re-analysis

The samples in this SDG did not require re-analysis.

Miscellaneous Information

Data Exception (DER) Documentation

The following DER was generated for this SDG: 1117440 1202730364 (CAMO-12-21734).

Additional Comments

Additional comments were not required for this SDG.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Method/Analysis Information

Product: Nitrate Nitrite by Cadmium Reduction
Analytical Batch: 1240584
Method: EPA 353.2 Nitrogen and Nitrate/Nitrite

Sample Analysis

The following samples were analyzed using the analytical protocol as established in EPA 353.2:

Sample ID	Client ID
309911002	CASA-12-21647
1202725425	Method Blank (MB)
1202725426	309548008(CAMO-12-21796) Sample Duplicate (DUP)
1202725429	309548008(CAMO-12-21796) Post Spike (PS)
1202725432	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-128 REV# 7.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Nutrient analysis was performed on a Lachat QuickChem FIA+ 8500 Series.

Continuing Calibration Blanks

All continuing calibration blanks (CCBs) associated with reported data from this batch were within acceptance limits.

Calibration Verification Information (CCV)

All continuing calibration verification standards (CCVs) associated with reported data from this batch were within acceptance limits.

Y Intercept Rule

The absolute value of the intercept is less than 3 times the MDL.

Quality Control (QC) Information

Method Blank (MB) Statement

The MB analyzed with this SDG met the acceptance criteria.

Laboratory Control Sample (LCS) Recovery

The LCS spike recovery met the acceptance limits.

Quality Control (QC) Designation

The following sample was selected for QC analysis: 309548008 (CAMO-12-21796).

Matrix Spike (MS)/Post Spike (PS) Recovery Statement

The MS/PS recoveries for this sample set were within the required acceptance limits.

Duplicate Relative Percent Difference (RPD) Statement

The RPD between the sample and its duplicate met the acceptance limits.

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

All samples in this SDG met the specified holding time.

Sample Preservation/Integrity

All the samples from this sample group met the preservation and integrity requirements of the method.

Sample Dilutions

The following sample in this sample group was diluted due to high concentration: 309911002 (CASA-12-21647).

Sample Re-analysis

The samples in this SDG did not require re-analysis.

Miscellaneous Information

Data Exception (DER) Documentation

Data exception reports (DERs) are generated to document procedural anomalies that may deviate from referenced SOP or contractual documents.

Additional Comments

Additional comments were not required for this SDG.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are

present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Method/Analysis Information

Product: Total Phosphorus
Analytical Batch: 1240581 **Method:** EPA 365.4 Phosphorus and Total in
Prep Batch : 1240580 **Method:** EEPA 365.4 Prep

Sample Analysis

The following samples were analyzed using the analytical protocol as established in EPA 365.4:

Sample ID	Client ID
309911002	CASA-12-21647
1202725412	Method Blank (MB)
1202725413	309454002(CAMO-12-21743) Sample Duplicate (DUP)
1202725414	309454002(CAMO-12-21743) Matrix Spike (MS)
1202725415	309454002(CAMO-12-21743) Matrix Spike Duplicate (MSD)
1202725416	Laboratory Control Sample (LCS)
1202730366	309454004(CAMO-12-21744) Sample Duplicate (DUP)
1202730367	309454004(CAMO-12-21744) Matrix Spike (MS)
1202730368	309454004(CAMO-12-21744) Matrix Spike Duplicate (MSD)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-103 REV# 8.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Nutrient analysis was performed on a Lachat QuickChem FIA+ 8000 Series.

Continuing Calibration Blanks

All continuing calibration blanks (CCBs) associated with reported data from this batch were within acceptance limits.

Calibration Verification Information (CCV)

All continuing calibration verification standards (CCVs) associated with reported data from this batch were within

acceptance limits.

Y Intercept Rule

The absolute value of the intercept is less than 3 times the MDL.

Quality Control (QC) Information

Method Blank (MB) Statement

The MB analyzed with this SDG met the acceptance criteria.

Laboratory Control Sample (LCS) Recovery

The LCS spike recovery met the acceptance limits.

Quality Control (QC) Designation

The following samples were selected for QC analysis: 309454002 (CAMO-12-21743) and 309454004 (CAMO-12-21744).

Matrix Spike (MS)/Post Spike (PS) Recovery Statement

The MS/PS recoveries for this sample set were within the required acceptance limits.

Matrix Spike Duplicate (MSD) Recovery Statement

The MSD recoveries for this sample set were within the required acceptance limits.

MS/MSD Relative Percent Difference (RPD) Statement

The RPDs between the spike and spike duplicate met the acceptance limits.

Duplicate Relative Percent Difference (RPD) Statement

One or more of the values for the sample and/or duplicate are less than 5 times the Practical Quantitation Limit (PQL), and the difference is within one PQL value; therefore, the RPD is not applicable. 1202730366 (CAMO-12-21744).

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

All samples in this SDG met the specified holding time.

Sample Preservation/Integrity

All the samples from this sample group met the preservation and integrity requirements of the method.

Sample Dilutions

The samples in this SDG did not require dilutions.

Sample Re-analysis

The samples in this SDG did not require re-analysis.

Miscellaneous Information

Data Exception (DER) Documentation

Data exception reports (DERs) are generated to document procedural anomalies that may deviate from referenced

SOP or contractual documents. A data exception report (DER) was not generated for this SDG.

Additional Comments

Additional comments were not required for this SDG.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Method/Analysis Information

Product: Solids, Total Dissolved
Analytical Batch: 1240523 **Method:** EPA 160.1 Solids and Dissolved-F

Sample Analysis

The following samples were analyzed using the analytical protocol as established in EPA 160.1:

Sample ID	Client ID
309911002	CASA-12-21647
1202725254	Method Blank (MB)
1202725255	309910002(CAMO-12-21742) Sample Duplicate (DUP)
1202725256	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-001 REV# 12.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Solids analysis was performed on a Sartorius Balance BAL216. Solids lab

Initial Calibration

All initial calibration requirements have been met for this SDG.

Quality Control (QC) Information

Method Blank (MB) Statement

The MB analyzed with this SDG met the acceptance criteria.

Laboratory Control Sample (LCS) Recovery

The LCS spike recovery met the acceptance limits.

Quality Control (QC) Designation

The following sample was selected for QC analysis: 309910002 (CAMO-12-21742).

Duplicate Relative Percent Difference (RPD) Statement

The Relative Percent Difference (RPD) between the sample and duplicate falls outside of the established acceptance limits because of the heterogeneous matrix of the sample: 1202725255 (CAMO-12-21742).

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

All samples in this SDG met the specified holding time.

Sample Re-analysis

The samples in this SDG did not require re-analysis.

Sample Aliquot

A sufficient amount of sample was provided by the client for analysis.

Miscellaneous Information

Data Exception (DER) Documentation

The following DER was generated for this SDG: 1115749 1202725255 (CAMO-12-21742).

Additional Comments

Additional comments were not required for this SDG.

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Method/Analysis Information

Product: Alkalinity
Analytical Batch: 1242061 **Method:** EPA 310.1 Total Alkalinity

Sample Analysis

The following samples were analyzed using the analytical protocol as established in EPA 310.1:

Sample ID	Client ID
309911002	CASA-12-21647
1202729069	Laboratory Control Sample (LCS)
1202729070	309996002(CASA-12-21775) Sample Duplicate (DUP)
1202729071	309996002(CASA-12-21775) Matrix Spike (MS)
1202729076	309907004(WTRO-12-23137) Sample Duplicate (DUP)
1202729077	309907004(WTRO-12-23137) Matrix Spike (MS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-GC-E-033 REV# 10.

Preparation/Analytical Method Verification

The SOP stated above has been prepared based on technical research and testing conducted by GEL Laboratories, LLC. and with guidance from the regulatory documents listed in this "Method/Analysis Information" section.

Calibration Information

The Titration analysis was performed on a Manually operated buret.

Initial Standardization

The titrant was properly standardized

Quality Control (QC) Information

Laboratory Control Sample (LCS) Recovery

The LCS spike recovery met the acceptance limits.

Quality Control (QC) Designation

The following samples were selected for QC analysis: 309907004 (WTRO-12-23137) and 309996002

(CASA-12-21775).

Matrix Spike (MS)/Post Spike (PS) Recovery Statement

The MS/PS recoveries for this sample set were within the required acceptance limits.

Duplicate Relative Percent Difference (RPD) Statement

The RPD between the sample and its duplicate met the acceptance limits.

Technical Information

GEL assigns holding times based on the date and time of sample collection. Those holding times expressed in hours are calculated in the AlphaLims system by hours. Those holding times expressed as days expire at midnight on the day of expiration.

Holding Times

All samples in this SDG met the specified holding time.

Sample Re-analysis

The samples in this SDG did not require re-analysis.

Miscellaneous Information

Data Exception (DER) Documentation

Data exception reports (DERs) are generated to document procedural anomalies that may deviate from referenced SOP or contractual documents. A data exception report (DER) was not generated for this SDG.

Additional Comments

Less volume was used because limited sample was provided by the client. 1202729076 (WTRO-12-23137) and 1202729077 (WTRO-12-23137).

Electronic Packaging Comment

This data package was generated using an electronic data processing program referred to as virtual packaging. In an effort to increase quality and efficiency, the laboratory has developed systems to generate all data packages electronically. The following change from traditional packages should be noted:

Analyst/peer reviewer initials and dates are not present on the electronic data files. Presently, all initials and dates are present on the original raw data. These hard copies are temporarily stored in the laboratory. The data validator will always sign and date the case narrative. Data that are not generated electronically, such as hand written pages, will be scanned and inserted into the electronic package.

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

Review Validation:

GEL requires all analytical data to be verified by a qualified data validator. In addition, all data designated for CLP or CLP-like packaging will receive a third level validation upon completion of the data package.

The following data validator verified the information presented in this case narrative:

Reviewer:  Date: 17Sep12

Sample Data Summary

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis Report for

ARSL001 ARS International (63641-10)

Client SDG: 12-1508 GEL Work Order: 309911

The Qualifiers in this report are defined as follows:

- * A quality control analyte recovery is outside of specified acceptance criteria
- ** Analyte is a surrogate compound
- H Analytical holding time was exceeded
- J Value is estimated
- U Analyte was analyzed for, but not detected above the MDL, MDA, or LOD.
- UI Gamma Spectroscopy--Uncertain identification

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless qualified on the Certificate of Analysis.

The designation ND, if present, appears in the result column when the analyte concentration is not detected above the detection limit.

This data report has been prepared and reviewed in accordance with GEL Laboratories LLC standard operating procedures. Please direct any questions to your Project Manager, Valerie Davis.

Reviewed by



GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Report Date: September 17, 2012

Company : Los Alamos National Laboratory
Address : PO Box 1663
TA-03, SM271, Drop Pt. 02U, Rm111
Los Alamos, New Mexico 87545

Contact: Keith Greene
Project: LANL-WQH Water Samples

Client SDG: 12-1508

Client Sample ID: CASA-12-21643

Project: ESHL00210

Sample ID: 309911001

Client ID: ARSL001

Matrix: W

Collect Date: 17-AUG-12 11:01

Receive Date: 21-AUG-12

Collector: Client

Parameter	Qualifier	Result	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
Carbon Analysis											
SW 9060 Total Organic Carbon "As Received"											
Total Organic Carbon Average	J	0.496	0.330	1.00	mg/L	1	TSM	08/29/12	1313	1241231	1
Nutrient Analysis											
Nitrogen, Total Kjeldahl (TKN) "As Received"											
Nitrogen, Total Kjeldahl	U	ND	0.035	0.100	mg/L	1	KLP1	09/05/12	1436	1240586	2

The following Prep Methods were performed:

Method	Description	Analyst	Date	Time	Prep Batch
EPA 351.2 Prep	EPA 351.2 Total Kjeldahl Nitrogen Prep	KLP1	09/04/12	1700	1240585

The following Analytical Methods were performed:

Method	Description	Analyst Comments
1	SW846 9060	
2	EPA 351.2	

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Report Date: September 17, 2012

Company : Los Alamos National Laboratory
 Address : PO Box 1663
 TA-03, SM271, Drop Pt. 02U, Rm111
 Los Alamos, New Mexico 87545

Contact: Keith Greene
 Project: LANL-WQH Water Samples

Client SDG: 12-1508

Client Sample ID: CASA-12-21647
 Sample ID: 309911002
 Matrix: W
 Collect Date: 17-AUG-12 11:01
 Receive Date: 21-AUG-12
 Collector: Client

Project: ESHL00210
 Client ID: ARSL001

Parameter	Qualifier	Result	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
Conductivity Analysis											
EPA120.1 Specific Conductivity "As Received"											
Conductivity		228	1.00	1.00	umhos/cm	1	TXT1	09/04/12	1533	1242995	1
Electrode Analysis											
EPA 150.1 pH "As Received"											
pH at Temp 19.4C	H	8.11	0.010	0.100	SU	1	LXA1	08/29/12	1435	1241830	2
Ion Chromatography											
EPA 300.0 Anions Liquid 28 day "As Received"											
Bromide	J	0.0882	0.067	0.200	mg/L	1	MAR1	09/01/12	0230	1241176	3
Chloride		4.83	0.067	0.200	mg/L	1					
Fluoride		0.378	0.033	0.100	mg/L	1					
Sulfate		12.0	0.133	0.400	mg/L	1					
Nutrient Analysis											
EPA 350.1 Nitrogen, Ammonia L "As Received"											
Nitrogen, Ammonia		0.124	0.017	0.050	mg/L	1	KLP1	08/27/12	1522	1240583	4
EPA 353.2 Nitrogen, Nitrate/Nitrite "As Received"											
Nitrogen, Nitrate/Nitrite		5.26	0.170	0.500	mg/L	10	KLP1	08/31/12	1027	1240584	5
EPA 365.4 Phosphorus, Total in "As Received"											
Phosphorus, Total as P	J	0.0349	0.017	0.050	mg/L	1	KLP1	09/05/12	1103	1240581	6
Solids Analysis											
EPA 160.1 Solids, Dissolved-F "As Received"											
Total Dissolved Solids		177	3.40	14.3	mg/L		LYG1	08/23/12	0948	1240523	7
Titration Analysis											
EPA 310.1 Total Alkalinity "As Received"											
Alkalinity, Total as CaCO3		69.7	0.725	1.00	mg/L		LXA1	08/29/12	1704	1242061	8
Carbonate alkalinity (CaCO3)	U	ND	0.725	1.00	mg/L						

The following Prep Methods were performed:

Method	Description	Analyst	Date	Time	Prep Batch
EPA 350.2 Prep	EPA 350.1 Ammonia Nitrogen Prep	KLP1	08/27/12	1330	1240582
EPA 365.4 Prep	EPA 365.4 Phosphorus, Total in liquid PR	KLP1	09/04/12	1700	1240580

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Report Date: September 17, 2012

Company : Los Alamos National Laboratory
Address : PO Box 1663
TA-03, SM271, Drop Pt. 02U, Rm111
Los Alamos, New Mexico 87545

Contact: Keith Greene
Project: LANL-WQH Water Samples

Client SDG: 12-1508

Client Sample ID: CASA-12-21647

Project: ESHL00210

Sample ID: 309911002

Client ID: ARSL001

Parameter	Qualifier	Result	DL	RL	Units	DF	Analyst	Date	Time	Batch	Method
-----------	-----------	--------	----	----	-------	----	---------	------	------	-------	--------

The following Analytical Methods were performed:

Method	Description	Analyst Comments
1	EPA 120.1	
2	EPA 150.1	
3	EPA 300.0	
4	EPA 350.1	
5	EPA 353.2	
6	EPA 365.4	
7	EPA 160.1	
8	EPA 310.1	

Quality Control Summary

GEL LABORATORIES LLC

2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Report Date: September 17, 2012
Page 1 of 5

Los Alamos National Laboratory
PO Box 1663
TA-03, SM271, Drop Pt. 02U, Rm111
Los Alamos, New Mexico
Contact: Keith Greene

Workorder: 309911

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time
Carbon Analysis											
Batch	1241231										
QC1202726978	309911001	DUP									
Total Organic Carbon Average	J	0.496	J	0.529	mg/L	6.44	^	(+/-1.00)	TSM	08/29/12	13:46
QC1202726979	310243006	DUP									
Total Organic Carbon Average	J	0.605	J	0.610	mg/L	0.823	^	(+/-1.00)		08/30/12	15:39
QC1202726982	LCS										
Total Organic Carbon Average	10.0			10.1	mg/L			101 (85%-115%)		08/29/12	12:31
QC1202726977	MB										
Total Organic Carbon Average			U	ND	mg/L					08/29/12	12:22
QC1202726980	309911001	PS									
Total Organic Carbon Average	10.0	J	0.496	9.97	mg/L			94.7 (65%-120%)		08/29/12	14:06
QC1202726981	310243006	PS									
Total Organic Carbon Average	10.0	J	0.605	8.47	mg/L			78.6 (65%-120%)		08/30/12	15:59
Conductivity Analysis											
Batch	1242995										
QC1202731308	309910002	DUP									
Conductivity			600	601	umhos/cm	0.167		(0%-10%)	TXT1	09/04/12	15:33
QC1202731309	LCS										
Conductivity	1410			1410	umhos/cm			99.7 (95%-105%)		09/04/12	15:29
Electrode Analysis											
Batch	1241830										
QC1202728424	309907006	DUP									
pH	H	6.43	H	6.44	SU	0.155		(0%-10%)	LXA1	08/29/12	14:23
QC1202728425	309910002	DUP									
pH	H	7.33	H	7.39	SU	0.815		(0%-10%)		08/29/12	14:33
QC1202728426	LCS										
pH	7.00			7.01	SU			100 (99%-101%)		08/29/12	14:15
Ion Chromatography											
Batch	1241176										
QC1202726830	310180003	DUP									
Bromide	U	ND	U	ND	mg/L	N/A			MAR1	09/01/12	05:47
Chloride		2.50		2.51	mg/L	0.336		(0%-20%)			
Fluoride		0.285		0.288	mg/L	0.977	^	(+/-0.100)			
Sulfate		3.83		3.79	mg/L	1.00		(0%-20%)			
QC1202726832	LCS										
Bromide	2.50			2.44	mg/L			97.5 (90%-110%)		09/01/12	01:57
Chloride	10.0			9.42	mg/L			94.2 (90%-110%)			
Fluoride	5.00			5.02	mg/L			100 (90%-110%)			
Sulfate	20.0			19.0	mg/L			95.2 (90%-110%)			
QC1202726829	MB										
Bromide			U	ND	mg/L					09/01/12	01:24
Chloride			U	ND	mg/L						

GEL LABORATORIES LLC

2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 309911

Page 2 of 5

Parname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time
Ion Chromatography											
Batch	1241176										
Fluoride			U	ND	mg/L						
Sulfate			U	ND	mg/L				MAR1	09/01/12	01:24
QC1202726831	310180003	PS									
Bromide	2.50	U	ND	2.50	mg/L		97.7	(90%-110%)		09/01/12	06:20
Chloride	10.0		2.50	12.2	mg/L		97.4	(90%-110%)			
Fluoride	5.00		0.285	5.14	mg/L		97.1	(90%-110%)			
Sulfate	20.0		3.83	23.2	mg/L		96.7	(90%-110%)			
Nutrient Analysis											
Batch	1240581										
QC1202725413	309454002	DUP									
Phosphorus, Total as P		J	0.0443	J	0.0428	mg/L	3.44	^	(+/-0.050)	KLP1	09/05/12 10:51
QC1202730366	309454004	DUP									
Phosphorus, Total as P			0.071	J	0.0419	mg/L	51.6	^	(+/-0.050)		09/05/12 10:55
QC1202725416	LCS										
Phosphorus, Total as P	1.00				1.11	mg/L		111	(84%-122%)		09/05/12 10:50
QC1202725412	MB										
Phosphorus, Total as P				J	0.0454	mg/L					09/05/12 10:49
QC1202725414	309454002	MS									
Phosphorus, Total as P	1.00	J	0.0443		1.07	mg/L		103	(46%-146%)		09/05/12 10:52
QC1202730367	309454004	MS									
Phosphorus, Total as P	1.00		0.071		1.13	mg/L		106	(46%-146%)		09/05/12 10:55
QC1202725415	309454002	MSD									
Phosphorus, Total as P	1.00	J	0.0443		1.10	mg/L	2.76	106	(0%-21%)		09/05/12 10:53
QC1202730368	309454004	MSD									
Phosphorus, Total as P	1.00		0.071		1.11	mg/L	1.79	104	(0%-21%)		09/05/12 10:56
Batch	1240583										
QC1202725418	309548008	DUP									
Nitrogen, Ammonia			0.0846		0.0644	mg/L	27.1	^	(+/-0.050)	KLP1	08/27/12 15:19
QC1202725424	LCS										
Nitrogen, Ammonia	1.00				1.10	mg/L		110	(90%-110%)		08/27/12 15:17
QC1202725417	MB										
Nitrogen, Ammonia				J	0.0359	mg/L					08/27/12 15:16
QC1202725420	309548008	MS									
Nitrogen, Ammonia	1.00		0.0846		1.09	mg/L		101	(90%-110%)		08/27/12 15:20
QC1202725422	309548008	MSD									
Nitrogen, Ammonia	1.00		0.0846		1.05	mg/L	3.74	96.5	(0%-15%)		08/27/12 15:20
Batch	1240584										
QC1202725426	309548008	DUP									
Nitrogen, Nitrate/Nitrite			0.441		0.442	mg/L	0.227		(0%-20%)	KLP1	08/31/12 10:23
QC1202725432	LCS										
Nitrogen, Nitrate/Nitrite	1.00				1.02	mg/L		102	(90%-110%)		08/31/12 10:21
QC1202725425	MB										
Nitrogen, Nitrate/Nitrite				U	ND	mg/L					08/31/12 10:20
QC1202725429	309548008	PS									
Nitrogen, Nitrate/Nitrite	1.00		0.441		1.44	mg/L		99.9	(90%-110%)		08/31/12 10:25
Batch	1240586										
QC1202725434	309791001	DUP									

GEL LABORATORIES LLC

2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 309911

Page 3 of 5

Parmname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time
Nutrient Analysis											
Batch	1240586										
Nitrogen, Total Kjeldahl		0.887		0.880	mg/L	0.792		(0%-20%)	KLP1	09/05/12	14:30
QC1202730363 309910001	DUP										
Nitrogen, Total Kjeldahl	J	0.0945	J	0.0605	mg/L	43.9 ^		(+/-0.100)		09/05/12	14:33
QC1202725437	LCS										
Nitrogen, Total Kjeldahl	1.00			0.923	mg/L		92.3	(90%-110%)		09/05/12	14:24
QC1202725433	MB										
Nitrogen, Total Kjeldahl			U	ND	mg/L					09/05/12	14:23
QC1202725435 309791001	MS										
Nitrogen, Total Kjeldahl	1.00	0.887		1.88	mg/L		99.3	(90%-110%)		09/05/12	14:31
QC1202730364 309910001	MS										
Nitrogen, Total Kjeldahl	1.00	J	0.0945	0.951	mg/L		85.7*	(90%-110%)		09/05/12	14:34
QC1202725436 309791001	MSD										
Nitrogen, Total Kjeldahl	1.00		0.887	1.93	mg/L	2.62	104	(0%-20%)		09/05/12	14:32
QC1202730365 309910001	MSD										
Nitrogen, Total Kjeldahl	1.00	J	0.0945	0.994	mg/L	4.42	90	(0%-20%)		09/05/12	14:35
Solids Analysis											
Batch	1240523										
QC1202725255 309910002	DUP										
Total Dissolved Solids		419		377	mg/L	10.4*		(0%-10%)	LYG1	08/23/12	09:48
QC1202725256	LCS										
Total Dissolved Solids	300			291	mg/L		97.1	(95%-105%)		08/23/12	09:48
QC1202725254	MB										
Total Dissolved Solids			U	ND	mg/L					08/23/12	09:48
Titration Analysis											
Batch	1242061										
QC1202729070 309996002	DUP										
Alkalinity, Total as CaCO3		88.1		88.1	mg/L	0.00		(0%-20%)	LXA1	08/29/12	17:52
Carbonate alkalinity (CaCO3)	U	ND	U	ND	mg/L	N/A					
QC1202729076 309907004	DUP										
Alkalinity, Total as CaCO3		25.2		25.2	mg/L	0.00		(0%-20%)		08/29/12	16:31
Carbonate alkalinity (CaCO3)	U	ND	U	ND	mg/L	N/A					
QC1202729069	LCS										
Alkalinity, Total as CaCO3	50.0			51.4	mg/L		103	(90%-110%)		08/29/12	15:33
QC1202729071 309996002	MS										
Alkalinity, Total as CaCO3	50.0	88.1		140	mg/L		105	(80%-120%)		08/29/12	17:54
QC1202729077 309907004	MS										
Alkalinity, Total as CaCO3	100	25.2		107	mg/L		81.8	(80%-120%)		08/29/12	16:34

Notes:

RER is calculated at the 95% confidence level (2-sigma).

The Qualifiers in this report are defined as follows:

- ** Analyte is a surrogate compound
- < Result is less than value reported
- > Result is greater than value reported
- A The TIC is a suspected aldol-condensation product
- B For General Chemistry and Organic analysis the target analyte was detected in the associated blank.

GEL LABORATORIES LLC

2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 309911

Page 4 of 5

Parname	NOM	Sample	Qual	QC	Units	RPD%	REC%	Range	Anlst	Date	Time
BD											
C											
D											
E											
E											
E											
F											
FB											
H											
J											
K											
L											
M											
M											
N											
N											
N/A											
N1											
ND											
NJ											
P											
Q											
R											
U											
UI											
UJ											
UL											
X											
Y											
Z											
^											
d											
h											

GEL LABORATORIES LLC

2040 Savage Road Charleston, SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 309911

Page 5 of 5

<u>Parmname</u>	<u>NOM</u>	<u>Sample</u>	<u>Qual</u>	<u>QC</u>	<u>Units</u>	<u>RPD%</u>	<u>REC%</u>	<u>Range</u>	<u>Anlst</u>	<u>Date</u>	<u>Time</u>
-----------------	------------	---------------	-------------	-----------	--------------	-------------	-------------	--------------	--------------	-------------	-------------

N/A indicates that spike recovery limits do not apply when sample concentration exceeds spike conc. by a factor of 4 or more.

^ The Relative Percent Difference (RPD) obtained from the sample duplicate (DUP) is evaluated against the acceptance criteria when the sample is greater than five times (5X) the contract required detection limit (RL). In cases where either the sample or duplicate value is less than 5X the RL, a control limit of +/- the RL is used to evaluate the DUP result.

* Indicates that a Quality Control parameter was not within specifications.

For PS, PSD, and SDILT results, the values listed are the measured amounts, not final concentrations.

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless qualified on the QC Summary.

Miscellaneous

DATA EXCEPTION REPORT

Mo.Day Yr. 30-AUG-12	Division: Industrial	Quality Criteria: Specifications	Type: Process
Instrument Type: BALANCE	Test / Method: EPA 160.1	Matrix Type: Liquid	Client Code: ESHL, FDAN, VVVC
Batch ID: 1240523	Sample Numbers: See below.		
Potentially affected work order(s)(SDG): 309866,309910(12-1509),309911(12-1508),309926,309996(12-1510)			
Application Issues: Failed RPD for DUP			
Specification and Requirements Exception Description:		DER Disposition:	
<p>1. Failed RPD for DUP: QC 1202725255DUP</p>		<p>1. The Relative Percent Difference (RPD) between the sample and duplicate falls outside of the established acceptance limits because of the heterogeneous matrix of the sample.</p>	

Originator's Name:

Lisa Gregory 30-AUG-12

Data Validator/Group Leader:

Julia Hamilton 30-AUG-12

DATA EXCEPTION REPORT

Mo.Day Yr. 05-SEP-12	Division: Industrial	Quality Criteria: Specifications	Type: Process
Instrument Type: LACHAT Flow Injection Analyzer	Test / Method: EPA 351.2	Matrix Type: Liquid	Client Code: ESHL
Batch ID: 1240586	Sample Numbers: See below.		
<p>Potentially affected work order(s)(SDG): 309455(12-1482),309791(12-1502),309910(12-1509),309911(12-1508),309996(12-1510),310070(12-1511),310346(12-1519),310348(12-1518)</p> <p>Application Issues:</p> <p>Failed Recovery for MS/PS</p>			
Specification and Requirements Exception Description:		DER Disposition:	
<p>1. Failed Recovery for MS:</p> <p>QC 1202730364MS</p>		<p>1. The spike recovery falls outside of the established acceptance limits. Since both the spike duplicate recovery and the RPD between the spike and spike duplicate fall within acceptance limits, the data is reported.</p>	

Originator's Name:

Kristen Parson 05-SEP-12

Data Validator/Group Leader:

Julia Hamilton 05-SEP-12

DATA EXCEPTION REPORT

Mo.Day Yr. 06-SEP-12	Division: Industrial	Quality Criteria: Specifications	Type: Process
Instrument Type: ELECTRODE	Test / Method: EPA 150.1	Matrix Type: Liquid	Client Code: CARE, DMAX, ESHL, FDAN
Batch ID: 1241830	Sample Numbers: See below.		

Potentially affected work order(s)(SDG): 309907(2012-2212),309908(2012-2210),309910(12-1509),309911(12-1508),309996(12-1510),310070(12-1511),310117,310178(EUI-8995),310180(12-1513),310197(2012-2231),310236

Application Issues:

Container scanning event for custody missed
Sample received out of holding

Specification and Requirements Exception Description:	DER Disposition:
<p>1. Sample received out of holding:</p> <p>309907 006</p> <p>309908 006,016</p> <p>309910 002</p> <p>309911 002</p> <p>309996 002,005</p> <p>310070 002,005</p> <p>310117 001</p> <p>310178 001</p> <p>310180 003,007,011,015,017</p> <p>310197 006,016,027</p> <p>310236 001</p> <p>2. Container scanning event for custody missed:</p> <p>310178001</p>	<p>1. Samples were received out of holding.</p> <p>2. Sample was not scanned to the analytical batch prior to analysis; however, it was in the analyst's custody at the time of analysis.</p>

Originator's Name:
Lindsey Jensen 06-SEP-12

Data Validator/Group Leader:
Julia Hamilton 06-SEP-12

Radiological Analysis

**Radiochemistry Case Narrative
ARS International (ARSL)
SDG 12-1508
Work Order 309911**

Method/Analysis Information

Product: Alphaspec Am241 Liquid
Analytical Method: DOE EML HASL-300, Am-05-RC Modified
Analytical Batch Number: 1239554

Sample ID	Client ID
309911001	CASA-12-21643
1202722699	Method Blank (MB)
1202722700	309870001(CAMO-12-21738) Sample Duplicate (DUP)
1202722701	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-RAD-A-011 REV# 22.

Calibration Information:

Calibration Information

All initial and continuing calibration requirements have been met. Calibrations are performed monthly using mixed alpha standards comprised of the following: Gd-148, Np-237, and Cm-244.

Standards Information

Standard solutions for these analysis are NIST traceable or verified with a NIST traceable standard and used before the expiration dates.

Sample Geometry

All counting sources were prepared in the same geometry as the calibration standards.

Quality Control (QC) Information:

Blank Information

Aliquots for samples 1202722699 (MB) and 1202722701 (LCS) were changed to 1.0 per client request.

Designated QC

The following sample was used for QC: 309870001 (CAMO-12-21738). The QC was from ARSL work order 309870.

QC Information

All of the QC samples met the required acceptance limits.

CSU

The blank result is less than 1.65 times the CSU.

Technical Information:

Holding Time

All sample procedures for this sample set were performed within the required holding time.

Sample Re-prep/Re-analysis

None of the samples in this sample set required reprep or reanalysis.

Miscellaneous Information:

Data Exception (DER) Documentation

Data exception reports are generated to document any procedural anomalies that may deviate from referenced SOP or contractual documents. A data exception report (DER) was not generated for this SDG.

Manual Integration

No manual integrations were performed on data in this batch.

Additional Comments

The MDCs are calculated using a blank population.

Blank Decision Level

The blank result is less than the decision level.

Qualifier Information

Manual qualifiers were not required.

Method/Analysis Information

Product: Alphaspec Pu, Liquid
Analytical Method: DOE EML HASL-300, Pu-11-RC Modified
Analytical Batch Number: 1239555

Sample ID	Client ID
309911001	CASA-12-21643
1202722702	Method Blank (MB)
1202722703	309870001(CAMO-12-21738) Sample Duplicate (DUP)
1202722704	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as

Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-RAD-A-011 REV# 22.

Calibration Information:

Calibration Information

All initial and continuing calibration requirements have been met. Calibrations are performed monthly using mixed alpha standards comprised of the following: Gd-148, Np-237, and Cm-244.

Standards Information

Standard solutions for these analysis are NIST traceable or verified with a NIST traceable standard and used before the expiration dates.

Sample Geometry

All counting sources were prepared in the same geometry as the calibration standards.

Quality Control (QC) Information:

Blank Information

Aliquots for samples 1202722702 (MB) and 1202722704 (LCS) were changed to 1.0 per client request.

Designated QC

The following sample was used for QC: 309870001 (CAMO-12-21738). The QC was from ARSL work order 309870.

QC Information

All of the QC samples met the required acceptance limits.

CSU

The blank result is less than 1.65 times the CSU.

Technical Information:

Holding Time

All sample procedures for this sample set were performed within the required holding time.

Sample Re-prep/Re-analysis

None of the samples in this sample set required reprep or reanalysis.

Miscellaneous Information:

Data Exception (DER) Documentation

Data exception reports are generated to document any procedural anomalies that may deviate from referenced SOP or contractual documents. A data exception report (DER) was not generated for this SDG.

Manual Integration

No manual integrations were performed on data in this batch.

Additional Comments

The MDCs are calculated using a blank population.

Blank Decision Level

The blank result is less than the decision level.

Qualifier Information

Manual qualifiers were not required.

Method/Analysis Information

Product: Alphaspec U, Liquid
Analytical Method: DOE EML HASL-300, U-02-RC Modified
Analytical Batch Number: 1239557

Sample ID	Client ID
309911001	CASA-12-21643
1202722705	Method Blank (MB)
1202722706	309870001(CAMO-12-21738) Sample Duplicate (DUP)
1202722707	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-RAD-A-011 REV# 22.

Calibration Information:

Calibration Information

All initial and continuing calibration requirements have been met. Calibrations are performed monthly using mixed alpha standards comprised of the following: Gd-148, Np-237, and Cm-244.

Standards Information

Standard solutions for these analysis are NIST traceable or verified with a NIST traceable standard and used before the expiration dates.

Sample Geometry

All counting sources were prepared in the same geometry as the calibration standards.

Quality Control (QC) Information:

Blank Information

Aliquots for samples 1202722705 (MB) and 1202722707 (LCS) were changed to 1.0 per client request.

Designated QC

The following sample was used for QC: 309870001 (CAMO-12-21738). The QC was from ARSL work order 309870.

QC Information

All of the QC samples met the required acceptance limits.

CSU

The blank result is less than 1.65 times the CSU.

Technical Information:

Holding Time

All sample procedures for this sample set were performed within the required holding time.

Sample Re-prep/Re-analysis

None of the samples in this sample set required reprep or reanalysis.

Miscellaneous Information:

Data Exception (DER) Documentation

Data exception reports are generated to document any procedural anomalies that may deviate from referenced SOP or contractual documents. A data exception report (DER) was not generated for this SDG.

Manual Integration

No manual integrations were performed on data in this batch.

Additional Comments

The MDCs are calculated using a blank population.

Blank Decision Level

The blank result is less than the decision level.

Qualifier Information

Manual qualifiers were not required.

Method/Analysis Information

Product:	GammaSpec
Analytical Method:	EPA 901.1
Analytical Batch Number:	1240464

Sample ID	Client ID
309911001	CASA-12-21643
1202725096	Method Blank (MB)
1202725097	309992001(WST03-12-23161) Sample Duplicate (DUP)
1202725098	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-RAD-A-013 REV# 24.

Calibration Information:

Calibration Information

All initial and continuing calibration requirements have been met. The initial Calibrations were performed in May 2012, June 2012 and July 2012.

Standards Information

Standard solutions for these analysis are NIST traceable or verified with a NIST traceable standard and used before the expiration dates.

Sample Geometry

All counting sources were prepared in the same geometry as the calibration standards.

Quality Control (QC) Information:

Blank Information

The blank volume is representative of the sample volume in this batch.

Designated QC

The following sample was used for QC: 309992001 (WST03-12-23161). The QC was from ARSL work order 309992.

QC Information

All of the QC samples met the required acceptance limits.

CSU

The blank result is less than 1.65 times the CSU.

Technical Information:

Holding Time

All sample procedures for this sample set were performed within the required holding time.

Sample Re-prep/Re-analysis

None of the samples in this sample set required reprep or reanalysis.

Miscellaneous Information:

Data Exception (DER) Documentation

Data exception reports are generated to document any procedural anomalies that may deviate from referenced SOP or contractual documents. A data exception report (DER) was not generated for this SDG.

Additional Comments

Additional comments were not required for this sample set.

Blank Decision Level

The blank result is less than the decision level.

Qualifier Information

Manual qualifiers were not required.

Method/Analysis Information

Product: GFPC, Sr90, liquid
Analytical Method: EPA 905.0 Modified
Analytical Batch Number: 1239939

Sample ID	Client ID
309911001	CASA-12-21643
1202723779	Method Blank (MB)
1202723780	309911001(CASA-12-21643) Sample Duplicate (DUP)
1202723781	309911001(CASA-12-21643) Matrix Spike (MS)
1202723782	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-RAD-A-004 REV# 16.

Calibration Information:

Calibration Information

All initial and continuing calibration requirements have been met. The initial Calibration was performed in April 2012.

Standards Information

Standard solutions for these analysis are NIST traceable or verified with a NIST traceable standard and used before the expiration dates.

Sample Geometry

All counting sources were prepared in the same geometry as the calibration standards.

Quality Control (QC) Information:

Blank Information

Aliquots for samples 1202723779 (MB) and 1202723782 (LCS) were changed to 1.0 per client request.

Designated QC

The following sample was used for QC: 309911001 (CASA-12-21643). The QC was from ARSL work order 309911.

QC Information

All of the QC samples met the required acceptance limits.

CSU

The blank result is less than 1.65 times the CSU.

Technical Information:

Holding Time

All sample procedures for this sample set were performed within the required holding time.

Sample Re-prep/Re-analysis

Samples 1202723780 (CASA-12-21643) and 309911001 (CASA-12-21643) were recounted due to high MDCs. The recounts are reported.

Chemical Recoveries

All chemical recoveries meet the required acceptance limits for this sample set.

Miscellaneous Information:**Data Exception (DER) Documentation**

Data exception reports are generated to document any procedural anomalies that may deviate from referenced SOP or contractual documents. A data exception report (DER) was not generated for this SDG.

Additional Comments

The matrix spike, 1202723781 (CASA-12-21643), aliquot was reduced to conserve sample volume.

Blank Decision Level

The blank result is less than the decision level.

Qualifier Information

Manual qualifiers were not required.

Method/Analysis Information

Product:	WSP-GrossA/B
Analytical Method:	EPA 900.0/SW846 9310
Analytical Batch Number:	1239941

Sample ID	Client ID
309911001	CASA-12-21643
1202723792	Method Blank (MB)
1202723793	309454003(CAMO-12-21736) Sample Duplicate (DUP)
1202723796	309454003(CAMO-12-21736) Matrix Spike (MS)
1202723797	309454003(CAMO-12-21736) Matrix Spike Duplicate (MSD)
1202723798	Laboratory Control Sample (LCS)

The samples in this SDG were analyzed on an "as received" basis.

SOP Reference

Procedure for preparation, analysis and reporting of analytical data are controlled by GEL Laboratories LLC as Standard Operating Procedure (SOP). The data discussed in this narrative has been analyzed in accordance with GL-RAD-A-001 REV# 15.

Calibration Information:**Calibration Information**

All initial and continuing calibration requirements have been met. The initial Calibration was performed in

September 2011. The discrimination settings are calibrated in beta discriminating mode to reduce beta to alpha crosstalk.

Standards Information

Standard solutions for these analysis are NIST traceable or verified with a NIST traceable standard and used before the expiration dates.

Sample Geometry

All counting sources were prepared in the same geometry as the calibration standards.

Quality Control (QC) Information:

Blank Information

Aliquots for samples 1202723792 (MB) and 1202723798 (LCS) were changed to 1.0 per client request.

Designated QC

The following sample was used for QC: 309454003 (CAMO-12-21736). The QC was from ARSL work order 309454.

QC Information

All of the QC samples met the required acceptance limits.

CSU

The blank result is less than 1.65 times the CSU.

Technical Information:

Holding Time

All sample procedures for this sample set were performed within the required holding time.

Sample Re-prep/Re-analysis

None of the samples in this sample set required reprep or reanalysis.

Chemical Recoveries

All chemical recoveries meet the required acceptance limits for this sample set.

Gross Alpha/Beta Preparation Information

High hygroscopic salt content in evaporated samples can cause the sample mass to fluctuate due to moisture absorption. To minimize this interference, the salts are converted to oxides by heating the sample under a flame until a dull red color is obtained. The conversion to oxides stabilizes the sample weight and ensures that proper alpha/beta efficiencies are assigned for each sample. Volatile radioisotopes of carbon, hydrogen, technetium, polonium and cesium may be lost during sample heating, especially to a dull red heat. For this sample set, the prepared planchet was counted for beta activity before being flamed. After flaming, the planchet was counted for alpha activity.

Miscellaneous Information:

Data Exception (DER) Documentation

Data exception reports are generated to document any procedural anomalies that may deviate from referenced SOP or contractual documents. A data exception report (DER) was not generated for this SDG.

Additional Comments

The matrix spike and matrix spike duplicate, 1202723796 (CAMO-12-21736) and 1202723797 (CAMO-12-21736), aliquots were reduced to conserve sample volume.

Blank Decision Level

The blank result is less than the decision level.

Qualifier Information

Manual qualifiers were not required.

Certification Statement

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless otherwise noted in the analytical case narrative.

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Qualifier Definition Report for

ARSL001 ARS International (63641-10)

Client SDG: 12-1508 GEL Work Order: 309911

The Qualifiers in this report are defined as follows:

- * A quality control analyte recovery is outside of specified acceptance criteria
- ** Analyte is a surrogate compound
- U Analyte was analyzed for, but not detected above the MDL, MDA, or LOD.
- UI Gamma Spectroscopy--Uncertain identification

Review/Validation

GEL requires all analytical data to be verified by a qualified data reviewer. In addition, all CLP-like deliverables receive a third level review of the fractional data package.

The following data validator verified the information presented in this data report:

Signature: 

Name: Heather McCarty

Date: 12 SEP 2012

Title: Analyst II

Sample Data Summary

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Company : Los Alamos National Laboratory
 Address : PO Box 1663
 TA-03, SM271, Drop Pt. 02U, Rm111
 Los Alamos, New Mexico 87545
 Contact: Keith Greene
 Project: LANL-WQH Water Samples

Report Date: September 12, 2012

Client Sample ID: CASA-12-21643
 Sample ID: 309911001
 Matrix: W
 Collect Date: 17-AUG-12
 Receive Date: 21-AUG-12
 Collector: Client

Project: ESHL00210
 Client ID: ARSL001

Parameter	Qualifier	Result	Uncertainty	DL	TPU	RL	Units	DF	Analyst	Date	Time	Batch	Mtd.
Rad Alpha Spec Analysis													
<i>Alphaspec Am241 Liquid "As Received"</i>													
Americium-241	U	0.00254	+/-0.00568	0.0348	+/-0.00568	0.050	pCi/L		JXH2	09/05/12	1400	1239554	1
<i>Alphaspec Pu, Liquid "As Received"</i>													
Plutonium-238	U	0.00	+/-0.00359	0.0171	+/-0.00359	0.050	pCi/L		JXH2	09/05/12	1401	1239555	2
Plutonium-239/240	U	0.00761	+/-0.00567	0.0306	+/-0.00568	0.050	pCi/L						
<i>Alphaspec U, Liquid "As Received"</i>													
Uranium-234		0.556	+/-0.0409	0.0735	+/-0.0549	1.00	pCi/L		JXH2	09/01/12	1814	1239557	3
Uranium-235/236	U	0.0136	+/-0.00833	0.0475	+/-0.00838	1.00	pCi/L						
Uranium-238		0.184	+/-0.0229	0.0373	+/-0.0258	0.500	pCi/L						
Rad Gamma Spec Analysis													
<i>Gammaspex "As Received"</i>													
Cesium-137	U	1.44	+/-1.86	7.55	+/-1.86	8.00	pCi/L		KXG3	08/29/12	1231	1240464	4
Cobalt-60	U	-2.91	+/-2.18	7.34	+/-2.18	8.00	pCi/L						
Neptunium-237	U	-6.01	+/-3.29	10.6	+/-3.29	10.0	pCi/L						
Potassium-40	U	-3.04	+/-17.0	76.7	+/-17.0	10.0	pCi/L						
Sodium-22	U	-0.956	+/-1.78	6.84	+/-1.78	10.0	pCi/L						
Rad Gas Flow Proportional Counting													
<i>GFPC, Sr90, liquid "As Received"</i>													
Strontium-90	U	0.149	+/-0.135	0.467	+/-0.135	0.500	pCi/L		VXC2	09/04/12	1420	1239939	5
<i>WSP-GrossA/B "As Received"</i>													
Beta	U	1.70	+/-0.823	2.66	+/-0.835	3.00	pCi/L		DYT1	08/31/12	0956	1239941	6
Alpha	U	-0.134	+/-0.376	2.15	+/-0.376	3.00	pCi/L		DYT1	09/01/12	1915	1239941	7

The following Analytical Methods were performed

Method	Description
1	DOE EML HASL-300, Am-05-RC Modified
2	DOE EML HASL-300, Pu-11-RC Modified
3	DOE EML HASL-300, U-02-RC Modified
4	EPA 901.1
5	EPA 905.0 Modified
6	EPA 900.0/SW846 9310
7	EPA 900.0/SW846 9310

Surrogate/Tracer Recovery	Test	Batch ID	Recovery%	Acceptable Limits
Americium-243 Tracer	Alphaspec Am241 Liquid "As Received"	1239554	88.4	(50%-105%)
Plutonium-242 Tracer	Alphaspec Pu, Liquid "As Received"	1239555	87.2	(50%-105%)
Uranium-232 Tracer	Alphaspec U, Liquid "As Received"	1239557	80.6	(50%-105%)
Strontium Carrier	GFPC, Sr90, liquid "As Received"	1239939	77.4	(50%-105%)

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

Certificate of Analysis

Company : Los Alamos National Laboratory
Address : PO Box 1663
TA-03, SM271, Drop Pt. 02U, Rm111
Los Alamos, New Mexico 87545

Report Date: September 12, 2012

Contact: Keith Greene
Project: LANL-WQH Water Samples

Client Sample ID: CASA-12-21643
Sample ID: 309911001

Project: ESHL00210
Client ID: ARSL001

Parameter	Qualifier	Result	Uncertainty	DL	TPU	RL	Units	DF	Analyst	Date	Time	Batch	Mtd.
Surrogate/Tracer	Recovery	Test								Batch ID	Recovery%	Acceptable Limits	

Notes:

TPU and Uncertainty are calculated at the 67% confidence level (1-sigma).

Quality Control Data

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Report Date: September 12, 2012
Page 1 of 6

Client : Los Alamos National Laboratory
PO Box 1663
TA-03, SM271, Drop Pt. 02U, Rm
Los Alamos, New Mexico
Contact: Keith Greene
Workorder: 309911

Parmname	NOM	Sample	Qual	QC	Units	RER	REC%	Range	Anlst	Date Time
Rad Alpha Spec										
Batch	1239554									
QC1202722700	309870001 DUP									
Americium-241	U	0.00	U	-0.00237	pCi/L	0.164		(0-1)	JXH2	09/05/1214:01
	Uncert:	+/-0.00314		+/-0.00411						
	TPU:	+/-0.00314		+/-0.00411						
**Americium-243 Tracer	2.67	2.61		2.45	pCi/L		91.6	(50%-105%)		
	Uncert:	+/-0.0772		+/-0.0797						
	TPU:	+/-0.135		+/-0.137						
QC1202722701	LCS									
Americium-241	1.42			1.42	pCi/L		100	(80%-120%)		09/05/1214:01
	Uncert:			+/-0.0541						
	TPU:			+/-0.0812						
**Americium-243 Tracer	2.14			1.78	pCi/L		83.4	(50%-105%)		
	Uncert:			+/-0.0663						
	TPU:			+/-0.113						
QC1202722699	MB									
Americium-241			U	-0.00193	pCi/L					09/05/1214:01
	Uncert:			+/-0.00335						
	TPU:			+/-0.00335						
**Americium-243 Tracer	2.14			1.92	pCi/L		90	(50%-105%)		
	Uncert:			+/-0.0644						
	TPU:			+/-0.111						
Batch	1239555									
QC1202722703	309870001 DUP									
Plutonium-238	U	0.00506	U	-0.00241	pCi/L	0.404		(0-1)	JXH2	09/05/1214:01
	Uncert:	+/-0.00506		+/-0.00418						
	TPU:	+/-0.00507		+/-0.00418						
Plutonium-239/240	U	0.00253	U	0.00965	pCi/L	0.345		(0-1)		
	Uncert:	+/-0.00438		+/-0.00591						
	TPU:	+/-0.00438		+/-0.00592						
**Plutonium-242 Tracer	2.41	2.08		2.16	pCi/L		89.4	(50%-105%)		
	Uncert:	+/-0.0783		+/-0.0764						
	TPU:	+/-0.131		+/-0.128						
QC1202722704	LCS									
Plutonium-238			U	0.00663	pCi/L			(80%-120%)		09/05/1214:16
	Uncert:			+/-0.00494						
	TPU:			+/-0.00495						
Plutonium-239/240	2.03			2.12	pCi/L		104	(80%-120%)		
	Uncert:			+/-0.0685						
	TPU:			+/-0.117						
**Plutonium-242 Tracer	1.93			1.60	pCi/L		82.8	(50%-105%)		
	Uncert:			+/-0.0658						
	TPU:			+/-0.108						
QC1202722702	MB									

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 309911

Page 2 of 6

Parmname	NOM	Sample Qual	QC	Units	RER	REC%	Range	Anlst	Date Time
Rad Alpha Spec									
Batch	1239555								
Plutonium-238		U	0.00404	pCi/L					
	Uncert:		+/-0.00404						
	TPU:		+/-0.00405						
Plutonium-239/240		U	-0.00202	pCi/L					
	Uncert:		+/-0.00452						
	TPU:		+/-0.00452						
**Plutonium-242 Tracer	1.93		1.63	pCi/L		84.4	(50%-105%)		
	Uncert:		+/-0.0626						
	TPU:		+/-0.104						
Batch	1239557								
QC1202722706	309870001 DUP								
Uranium-234		0.446	0.379	pCi/L	0.372		(0-1)	JXH2	09/01/1218:14
	Uncert:	+/-0.0373	+/-0.0335						
	TPU:	+/-0.0474	+/-0.0416						
Uranium-235/236		U	0.0267	U	0.00	0.621	(0-1)		
	Uncert:	+/-0.0106	+/-0.0108						
	TPU:	+/-0.0107	+/-0.0108						
Uranium-238		0.203	0.182	pCi/L	0.199		(0-1)		
	Uncert:	+/-0.0237	+/-0.0214						
	TPU:	+/-0.0271	+/-0.0243						
**Uranium-232 Tracer	2.74	2.34	2.44	pCi/L		89	(50%-105%)		
	Uncert:	+/-0.0865	+/-0.0804						
	TPU:	+/-0.197	+/-0.192						
QC1202722707	LCS								
Uranium-234			2.40	pCi/L					09/01/1218:14
	Uncert:		+/-0.0807						
	TPU:		+/-0.179						
Uranium-235/236			0.0924	pCi/L					
	Uncert:		+/-0.0193						
	TPU:		+/-0.0202						
Uranium-238	2.67		2.67	pCi/L		100	(80%-120%)		
	Uncert:		+/-0.0848						
	TPU:		+/-0.197						
**Uranium-232 Tracer	2.19		1.39	pCi/L		63.3	(50%-105%)		
	Uncert:		+/-0.0772						
	TPU:		+/-0.165						
QC1202722705	MB								
Uranium-234		U	-0.0191	pCi/L					09/01/1218:14
	Uncert:		+/-0.00978						
	TPU:		+/-0.00978						
Uranium-235/236		U	0.00	pCi/L					
	Uncert:		+/-0.00541						
	TPU:		+/-0.00541						
Uranium-238		U	0.00309	pCi/L					
	Uncert:		+/-0.00818						
	TPU:		+/-0.00819						
**Uranium-232 Tracer	2.19		1.24	pCi/L		56.8	(50%-105%)		
	Uncert:		+/-0.0829						

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 309911

Page 3 of 6

Parmname	NOM	Sample Qual	QC	Units	RER	REC%	Range	Anlst	Date Time
Rad Alpha Spec									
Batch	1239557								
		TPU:							+/-0.170
Rad Gamma Spec									
Batch	1240464								
QC1202725097	309992001 DUP								
Cesium-137		6.43	U	5.89	pCi/L	0.0688	(0-1)	KXG3	08/30/1213:06
		Uncert:		+/-1.80					
		TPU:		+/-1.80					
Cobalt-60		U 0.229	U	-1.18	pCi/L	0.251	(0-1)		
		Uncert:		+/-1.70					
		TPU:		+/-1.70					
Neptunium-237		U 0.125	U	2.65	pCi/L	0.220	(0-1)		
		Uncert:		+/-3.18					
		TPU:		+/-3.18					
Potassium-40		72.2	U	24.3	pCi/L	0.480	(0-1)		
		Uncert:		+/-26.2					
		TPU:		+/-26.2					
Sodium-22		U -0.0135	U	-0.66	pCi/L	0.121	(0-1)		
		Uncert:		+/-1.53					
		TPU:		+/-1.53					
QC1202725098	LCS								
Americium-241	2780			2880	pCi/L	103	(80%-120%)		08/30/1211:45
		Uncert:		+/-224					
		TPU:		+/-224					
Cesium-137	6120			6180	pCi/L	101	(80%-120%)		
		Uncert:		+/-283					
		TPU:		+/-283					
Cobalt-60	5820			5820	pCi/L	100	(80%-120%)		
		Uncert:		+/-247					
		TPU:		+/-247					
Neptunium-237			U	1.53	pCi/L				
		Uncert:		+/-25.8					
		TPU:		+/-25.8					
Potassium-40			U	31.8	pCi/L				
		Uncert:		+/-52.9					
		TPU:		+/-52.9					
Sodium-22			U	-10.9	pCi/L				
		Uncert:		+/-8.78					
		TPU:		+/-8.78					
QC1202725096	MB								
Cesium-137			U	0.137	pCi/L				08/30/1211:39
		Uncert:		+/-1.27					
		TPU:		+/-1.27					
Cobalt-60			U	1.65	pCi/L				
		Uncert:		+/-1.50					
		TPU:		+/-1.50					
Neptunium-237			U	1.72	pCi/L				
		Uncert:		+/-2.77					
		TPU:		+/-2.77					

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 309911

Page 4 of 6

Parmname	NOM	Sample	Qual	QC	Units	RER	REC%	Range	Anlst	Date Time
Rad Gamma Spec										
Batch	1240464									
Potassium-40			U	-20.3	pCi/L					
				Uncert: +/-18.1						
				TPU: +/-18.1						
Sodium-22			U	2.05	pCi/L					
				Uncert: +/-1.65						
				TPU: +/-1.65						
Rad Gas Flow										
Batch	1239939									
QC1202723780	309911001 DUP									
Strontium-90	U	0.149	U	0.293	pCi/L	0.251		(0-1)	VXC2	09/05/1213:07
		Uncert: +/-0.135		+/-0.149						
		TPU: +/-0.135		+/-0.151						
**Strontium Carrier	8.14	6.30		6.80	mg		83.5	(50%-105%)		
QC1202723782	LCS									
Strontium-90		25.0		27.5	pCi/L		110	(80%-120%)		08/30/1217:07
				Uncert: +/-0.613						
				TPU: +/-2.37						
**Strontium Carrier	8.14			6.70	mg		82.3	(50%-105%)		
QC1202723779	MB									
Strontium-90			U	-0.0406	pCi/L					08/30/1217:06
				Uncert: +/-0.0673						
				TPU: +/-0.0673						
**Strontium Carrier	8.14			6.70	mg		82.3	(50%-105%)		
QC1202723781	309911001 MS									
Strontium-90	125	U	0.149	130	pCi/L		104	(75%-125%)		08/30/1217:07
				Uncert: +/-0.135						
				TPU: +/-0.135						
**Strontium Carrier	8.14			6.90	mg		84.8	(50%-105%)		
Batch	1239941									
QC1202723793	309454003 DUP									
Alpha	U	-0.106	U	0.181	pCi/L	0.177		(0-1)	DYT1	09/01/1219:27
		Uncert: +/-0.385		+/-0.423						
		TPU: +/-0.386		+/-0.424						
Beta	U	1.37	U	1.70	pCi/L	0.129		(0-1)		08/31/1209:35
		Uncert: +/-0.619		+/-0.637						
		TPU: +/-0.630		+/-0.652						
QC1202723798	LCS									
Alpha	12.0			11.2	pCi/L		92.9	(80%-120%)		09/01/1219:17
				Uncert: +/-0.627						
				TPU: +/-1.23						
Beta	49.9			47.7	pCi/L		95.5	(80%-120%)		08/31/1209:48
				Uncert: +/-0.865						
				TPU: +/-4.03						
QC1202723792	MB									
Alpha			U	0.0052	pCi/L					09/01/1219:17
				Uncert: +/-0.0545						
				TPU: +/-0.0546						
Beta			U	0.189	pCi/L					08/31/1209:33

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 309911

Page 5 of 6

Parmname	NOM	Sample Qual	QC	Units	RER	REC%	Range	Anlst	Date Time
Rad Gas Flow									
Batch	1239941								
		Uncert:					+/-0.180		
		TPU:					+/-0.180		
QC1202723796	309454003	MS							
Alpha	481	U	-0.106	448	pCi/L	93	(75%-125%)		09/01/1219:17
		Uncert:	+/-0.385				+/-25.5		
		TPU:	+/-0.386				+/-45.6		
Beta	2000	U	1.37	2020	pCi/L	101	(75%-125%)		08/31/1209:48
		Uncert:	+/-0.619				+/-37.0		
		TPU:	+/-0.630				+/-171		
QC1202723797	309454003	MSD							
Alpha	481	U	-0.106	456	pCi/L	0.0453	94.7	(0-1)	09/01/1219:17
		Uncert:	+/-0.385				+/-26.1		
		TPU:	+/-0.386				+/-46.6		
Beta	2000	U	1.37	2140	pCi/L	0.174	107	(0-1)	08/31/1209:48
		Uncert:	+/-0.619				+/-37.6		
		TPU:	+/-0.630				+/-181		

Notes:

The Qualifiers in this report are defined as follows:

- ** Analyte is a surrogate compound
- < Result is less than value reported
- > Result is greater than value reported
- A The TIC is a suspected aldol-condensation product
- B For General Chemistry and Organic analysis the target analyte was detected in the associated blank.
- BD Results are either below the MDC or tracer recovery is low
- C Analyte has been confirmed by GC/MS analysis
- D Results are reported from a diluted aliquot of the sample
- E General Chemistry--Concentration of the target analyte exceeds the instrument calibration range
- E Metals--%difference of sample and SD is >10%. Sample concentration must meet flagging criteria
- E Organics--Concentration of the target analyte exceeds the instrument calibration range
- F Estimated Value
- FB Mercury was found present at quantifiable concentrations in field blanks received with these samples. Data associated with the blank are deemed invalid for reporting to regulatory agencies
- H Analytical holding time was exceeded
- J Value is estimated
- K Analyte present. Reported value may be biased high. Actual value is expected to be lower.
- L Analyte present. Reported value may be biased low. Actual value is expected to be higher.
- M M if above MDC and less than LLD
- M Matrix Related Failure
- N Metals--The Matrix spike sample recovery is not within specified control limits
- N Organics--Presumptive evidence based on mass spectral library search to make a tentative identification of the analyte (TIC). Quantitation is based on nearest internal standard response factor
- N/A RPD or %Recovery limits do not apply.
- N1 See case narrative

GEL LABORATORIES LLC

2040 Savage Road Charleston SC 29407 - (843) 556-8171 - www.gel.com

QC Summary

Workorder: 309911

Page 6 of 6

Parname	NOM	Sample Qual	QC	Units	RER	REC%	Range	Anlst	Date	Time
ND	Analyte concentration is not detected above the detection limit									
NJ	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier									
P	Organics--The concentrations between the primary and confirmation columns/detectors is >40% different. For HPLC, difference is also <70%									
Q	One or more quality control criteria have not been met. Refer to the applicable narrative or DER.									
R	Sample results are rejected									
U	Analyte was analyzed for, but not detected above the MDL, MDA, or LOD.									
UI	Gamma Spectroscopy--Uncertain identification									
UJ	Gamma Spectroscopy--Uncertain identification									
UL	Not considered detected. The associated number is the reported concentration, which may be inaccurate due to a low bias.									
X	Consult Case Narrative, Data Summary package, or Project Manager concerning this qualifier									
Y	QC Samples were not spiked with this compound									
Z	Paint Filter Test--Particulates passed through the filter, however no free liquids were observed.									
^	RPD of sample and duplicate evaluated using +/-RL. Concentrations are <5X the RL. Qualifier Not Applicable for Radiochemistry.									
d	5-day BOD--The 2:1 depletion requirement was not met for this sample									
h	Preparation or preservation holding time was exceeded									

N/A indicates that spike recovery limits do not apply when sample concentration exceeds spike conc. by a factor of 4 or more.

** Indicates analyte is a surrogate/tracer compound.

^ The Relative Percent Difference (RPD) obtained from the sample duplicate (DUP) is evaluated against the acceptance criteria when the sample is greater than five times (5X) the contract required detection limit (RL). In cases where either the sample or duplicate value is less than 5X the RL, a control limit of +/- the RL is used to evaluate the DUP result.

For PS, PSD, and SDILT results, the values listed are the measured amounts, not final concentrations.

Where the analytical method has been performed under NELAP certification, the analysis has met all of the requirements of the NELAC standard unless qualified on the QC Summary.